

# Standard Test Method for Nondestructive Assay of Plutonium, Tritium and <sup>241</sup>Am by Calorimetric Assay<sup>1</sup>

This standard is issued under the fixed designation C1458; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\varepsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method describes the nondestructive assay (NDA) of plutonium, tritium, and  $^{241}$ Am using heat flow calorimetry. For plutonium the typical range of applicability, depending on the isotopic composition, corresponds to ~0.1 g to ~5 g quantities while for tritium the typical range extends from ~0.001 g to ~400 g. This test method can be applied to materials in a wide range of container sizes up to 380 L. It has been used routinely to assay items whose thermal power ranges from 0.001 W to 135 W.

1.2 This test method requires knowledge of the relative abundances of the plutonium isotopes and the  $^{241}$ Am/Pu mass ratio to determine the total plutonium mass.

1.3 This test method provides a direct measure of tritium content.

1.4 This test method provides a measure of  $^{241}$ Am either as a single isotope or mixed with plutonium.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

# 2. Referenced Documents

2.1 ASTM Standards:<sup>2</sup>

C697 Test Methods for Chemical, Mass Spectrometric, and Spectrochemical Analysis of Nuclear-Grade Plutonium Dioxide Powders and Pellets

- C1009 Guide for Establishing and Maintaining a Quality Assurance Program for Analytical Laboratories Within the Nuclear Industry
- C1030 Test Method for Determination of Plutonium Isotopic Composition by Gamma-Ray Spectrometry
- C1592 Guide for Nondestructive Assay Measurements

C1673 Terminology of C26.10 Nondestructive Assay Methods

2.2 ANSI Standard:<sup>3</sup>

ANSI N15.36 Measurement Control Program – Nondestructive Assay Measurement Control and Assurance

## 3. Terminology

3.1 Definitions:

3.1.1 Terms shall be defined in accordance with C26.10 Terminology C1673 except for the following:

3.1.2 *active mode*—a mode of calorimeter operation where an external power, applied by means of a series of band heaters for instance, is used to maintain the calorimeter at constant temperature.

3.1.3 *baseline*, *n*—the calorimeter output signal with no heat-generating item in the calorimeter item chamber.

3.1.4 *base power,* n—a constant thermal power applied in a calorimeter through an electrical resistance heater with no heat-generating item in the item chamber.

3.1.5 *equilibrium*, *n*—the point at which the temperature of the calorimeter measurement cell and the item being measured stabilizes.

3.1.6 *heat distribution error, n*—the bias arising from the location of the heat source within the calorimeter chamber.

3.1.7 *passive mode*, *n*—a mode of calorimeter operation where temperature change (caused by the heat to be quantified) is measured by means of temperature sensors with no external power applied except in the case of Wheatstone bridge temperature sensors where electrical current is needed to excite the bridge circuit.

<sup>&</sup>lt;sup>1</sup> This test method is under the jurisdiction of ASTM Committee C26 on Nuclear Fuel Cycle and is the direct responsibility of Subcommittee C26.10 on Non Destructive Assay.

Current edition approved March 1, 2016. Published April 2016. Originally approved in 2000. Last previous edition approved in 2009 as  $C1458 - 09^{\varepsilon 1}$ . DOI: 10.1520/C1458-16.

<sup>&</sup>lt;sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>&</sup>lt;sup>3</sup> Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, http://www.ansi.org.

3.1.8 *power compensation mode, n*—a mode of calorimeter operation where a constant applied thermal power is maintained in a calorimeter measurement chamber through the use of an electric resistance heater in a closed loop control system. Historically, Mound Laboratory used to call this "Servo Control."

3.1.9 *sensitivity, n*—the change in calorimeter response per Watt of thermal power (usually in units of  $\mu$ V/W) for a heat flow calorimeter.

3.1.10 *specific power,* n—the rate of energy emission by ionizing radiation per unit mass of a radionuclide, such as <sup>241</sup>Am or tritium.

3.1.11 *thermal diffusivity, n*—the ratio of thermal conductivity to the heat capacity. It measures the ability of a material to conduct thermal energy relative to its ability to store thermal energy.

3.1.12 *thermal power*, *n*—the rate at which heat is generated in a radioactively decaying item.

3.1.13 *thermal resistance*, *n*—ratio of the temperature difference at two different surfaces to the heat flux through the surfaces at equilibrium.

3.1.14 *thermal time constant*, *n*—an exponential decay constant describing the rate at which a temperature approaches a constant value. The combination of the item and its container will have numerous thermal time constants.

#### 4. Summary of Test Method

4.1 The item is placed in the calorimeter measurement chamber and the heat flow at equilibrium, that is, the thermal power, from the item is determined by temperature sensors and associated electronic equipment.

4.2 The thermal power emitted by a test item is directly related to the quantity of the radioactive material in it. The power generated by ionizing radiation absorbed in the item is measured by the calorimeter.

4.3 The total power  $W_{item}$  from a mixture of isotopes in the test item is the sum of the power from each heat-producing isotope:

$$W_{item} = \Sigma m_i P_i \tag{1}$$

where  $m_i$  is the mass of the *i*<sup>th</sup> isotope and  $P_i$  is the specific power (W/g isotope) with the sum taken over all heatproducing isotopes, most usually <sup>238</sup>Pu, <sup>239</sup>Pu, <sup>240</sup>Pu, <sup>241</sup>Pu, <sup>242</sup>Pu, and <sup>241</sup>Am for Pu-bearing items

4.4 The mass of Pu, tritium, or  $^{241}$ Am is calculated from the measured thermal power of an item, refer to 11.3.

4.4.1 When tritium is the only heat source the measured thermal power can be directly converted into the tritium mass using the specific power of tritium,  $P_{eff} = (0.3240 \pm 0.00045)$  (SD) W/g (1).<sup>4</sup>

4.4.2 When <sup>241</sup>Am is the only heat source, as a single isotope, the measured thermal power can be directly converted into mass using the specific power of <sup>241</sup>Am,  $P_{eff} = (0.1142 \pm 0.00042)$  (SD) W/g (see Table 1).

4.4.3 The <sup>241</sup>Am mass,  $m_{Am}$ , in a plutonium-bearing item is determined by multiplying the Pu mass by the Am/Pu mass ratio. The Am/Pu mass ratio is typically determined by gamma-ray spectroscopy.

## 5. Significance and Use

5.1 This test method is considered to be the most accurate NDA technique for the assay of many physical forms of Pu. Isotopic measurements by gamma-ray spectroscopy or destructive analysis techniques are part of this test method when it is applied to the assay of Pu.

5.1.1 Calorimetry has been applied to a wide variety of Pu-bearing solids including metals, alloys, oxides, fluorides, mixed Pu-U oxides, mixed oxide fuel pins, waste, and scrap, for example, ash, ash heels, salts, crucibles, and graphite scarfings) (2, 3). This test method has been routinely used at U.S. and European facilities for Pu process measurements and nuclear material accountability since the mid 1960's (2-9).

5.1.2 Pu-bearing materials have been measured in calorimeter containers ranging in size from about 0.025 m to about 0.63 m in diameter and from about 0.076 m to about 1.38 m in height.

5.1.3 Gamma-ray spectroscopy typically is used to determine the Pu isotopic composition and <sup>241</sup>Am to Pu ratio (see Test Method C1030). However, isotopic information from mass spectrometry and alpha counting measurements may be used instead (see Test Method C697).

5.2 This test method is considered to be the most accurate NDA method for the measurement of tritium. For many physical forms of tritium compounds calorimetry is currently the only practical measurement technique available.

5.3 Physical standards representative of the materials being assayed are not required for the test method.

5.3.1 This test method is largely independent of the elemental distribution of the nuclear materials in the matrix.

5.3.2 The accuracy of the method can be degraded for materials with inhomogeneous isotopic composition.

5.4 The thermal power measurement is traceable to national measurement systems through electrical standards used to directly calibrate the calorimeters or to calibrate secondary  $^{238}$ Pu heat standards.

5.5 Heat-flow calorimetry has been used to prepare secondary standards for neutron and gamma-ray assay systems (7-12).

5.6 Four parameters of the item and the item packaging affect measurement time. These four parameters are density, mass, thermal conductivity, and change in temperature. The measurement well of passive calorimeters will also affect measurement time because it too will need to come to the new equilibrium temperature. Calorimeters operated in power compensation mode maintain a constant measurement well temperature and have no additional effect on measurement time.

5.6.1 Calorimeter measurement times range from 20 minutes (13) for smaller, temperature-conditioned containers up to 72 h (14) for larger containers and items with long thermaltime constants.

5.6.2 Measurement times may be reduced by using equilibrium prediction techniques, by temperature preconditioning of

<sup>&</sup>lt;sup>4</sup> The boldface numbers in parentheses refer to the list of references at the end of this standard.

the item to be measured, by operating the calorimeter using the power compensation technique, or by optimization of the item container (low thermal mass and high thermal conductivity) and packaging.

## 6. Interferences

6.1 For plutonium-bearing items alpha decay heating is usually the dominant heat source and any interferences are typically negligible. These minor effects included the points 6.2 and 6.3.

6.2 Interferences for calorimetry are those processes that would add or subtract thermal power from the power of the radionuclides being assayed. Some examples include phase changes, endothermic or exothermic chemical reactions, such as oxidation, radiolysis of liquids, and bacterial action.

6.3 For the  $P_{eff}$  calculation, charged particles are assumed to be totally absorbed in the matrix. The contribution of high energy gamma-rays (for example, 1.173 MeV and 1.333 MeV of the <sup>60</sup>Co) needs to be calculated to not underestimate their contribution which could bias the measurement results. These assessments requiring subject matter expertise are usually based on high-resolution gamma spectroscopy (HRGS) and need to be considered in a case by case basis. These effects are usually considered negligible (**15**, **16**).

6.4 The loss of energy by escaping neutrons following spontaneous and induced fissions is about 1 %. The loss of energy by escaping neutrons following ( $\alpha$ ,  $\eta$ ) reactions is around 0.007 % per alpha particle. These effects are also usually considered negligible (**15**, **16**).

## 7. Apparatus

7.1 Calorimeters are designed to measure different sizes and quantities of nuclear material. Different types of heat-flow calorimeter systems share the common attributes listed below.

7.1.1 *Measurement Chamber*—Heat flow calorimeters typically have a cylindrical, cuboid, or hexagonal measurement chamber from which all of the heat flow generated by radioactive decay is directed through temperature sensors. It may also have a reference chamber in which a dummy can is placed (so-called twin cell calorimeter) (15, 17, 18).

7.1.1.1 An electrical heater may be built into the walls or the base of the chamber to provide measured amounts of thermal power into the calorimeter well.

7.1.1.2 Insulation or active heaters (or both) should surround the cavity to shield the chamber from outside temperature variations that would influence the thermal power measurement. Typically, an insulated plug or a temperature-controlled plug with an independent electrical heater is inserted above the item container inside the calorimeter. For some calorimeter types an insulating plug is installed permanently below the measurement chamber.

7.1.2 *Calorimeter Can*—The item to be measured may be placed in a special can that is designed to be inserted and removed easily from the calorimeter. It will typically have only a small air gap to provide good thermal conductivity between the outer surface of the can and the inner surface of the measurement chamber.

7.1.3 *Temperature Sensors*—Temperature sensors consist of thermistors, thermocouples, temperature sensitive resistance wires, thermopiles, or Peltier modules working in Seebeck mode. Temperature sensors should be placed as close as possible in a geometrical homogeneous network all around the can.

7.1.4 *Thermal Sink*—The temperature increases due to heat flows generated by items are measured against a reference temperature of a thermal sink. The thermal sink could be a water bath, air bath, a solid block, usually metallic block, maintained at a constant temperature or a combination of a solid block and water bath or air bath.

7.1.5 *Electrical Components*—Sensitive, stable electronic components are required for accurate calorimeter measurements.

7.1.5.1 High precision voltmeters or voltage measurement devices are required to quantify the voltage changes generated from the temperature sensors. The resolution should be better than one part per million of the voltage range.

7.1.5.2 Stable power supplies are necessary to provide constant current to Wheatstone bridge sensors and calorimeter heaters.

7.1.5.3 Precision resistors with certified resistances traceable to a national measurement system may be used with calibrated voltmeters to accurately determine electrical power delivered to heaters in the calorimeter chamber. If radioactive heat standards are used as part of the measurement control program the calorimeter voltmeters need not be calibrated nor are precision resistors required.

7.1.5.4 For a calorimeter operated in the power compensation mode digital-to-analog controller units are used to supply power to an internal resistance heater to maintain a constant temperature differential across thermal resistances.

7.1.6 *Heat Standards*—Thermal power standards are required to calibrate the calorimeter and may be used as measurement control standards to check the stability of calorimeter performance (**19-22**).

7.1.6.1 Radioactive heat standards, typically <sup>238</sup>Pu heat sources, also may be used to calibrate calorimeters over a range of thermal powers. These standards are calibrated against electrical standards traceable to a national measurement system. The certified power is typically decay corrected to the nearest day using certified decay tables.

7.1.6.2 Removable electrical heaters may be used to calibrate calorimeters. For this type of standard the power generated by the heater must be measured with electrical equipment regularly calibrated against standards or standard methods traceable to a national measurement system. The power supplied to the electrical calibration heater may be varied over the calibration range.

7.1.7 *Wheatstone Bridge*—When temperature sensitive resistance wire is used as the sensor, it is arranged in a Wheatstone bridge configuration shown in Fig. 1.

7.1.8 *Data Acquisition System*—Calorimeter data collection is performed using computer-based data acquisition systems. The system should be able to read signal voltages or resistances at a fixed time frequency and be able to calculate and report a

🖽 C1458 – 16



FIG. 1 Calorimeter Wheatstone Bridge Circuit

power value from the item using software that detects equilibrium. Graphics and numerical data indicating system power and temperatures may be displayed to aid the operator.

7.1.9 *Adapters*—Low mass cylindrical metal adapters may be fabricated to accept smaller calorimeter containers in the calorimeter well, and thus, provide good thermal contact between the outer container surface and calorimeter inner wall. Heat-conducting metal foil or metal gauze fill material, typically Al or Cu, or metal shot can be used in place of machined metal adapters. Smaller items may be placed in the calorimeter container and the void space inside the container may be filled with metal fill material or shot to provide good thermal contact. Lack of fill material does not preclude measurement of the item but it will increase the time to reach the calorimetric equilibrium and thus increase the total measurement time.

7.1.10 *Loading Apparatus*—A hoist or other assist may be used to load and unload items. Robotic loading systems may be used to handle the items.

#### 8. Heat-Flow Calorimeter Systems

8.1 A heat flow calorimeter system consists of an item chamber thermally insulated from a constant temperature environment by a thermal resistance. It may also include a reference chamber (twin calorimeter) also insulated from a constant temperature environment by a thermal resistance (15, 17, 18). When an item is placed in the calorimeter the temperature difference across the thermal resistance is disturbed and the difference changes with time until it converges to a constant value and equilibrium is achieved. The magnitude of the shift in the measured voltage (passive mode) or supplied power (power compensation mode) is used to determine the thermal power of the item in the calorimeter.

8.1.1 The curve of the temperature difference displaying the approach to the calorimetric equilibrium is a function of several exponentials with different time constants related to the specific heats and thermal conductivities of the item matrix material, packaging, and the calorimeter.

8.1.1.1 Equilibrium may be detected by visual inspection of the measurement data versus time or through statistical tests performed on a set of the latest data points in the time series.

8.1.1.2 Statistical prediction algorithms may be used during transient temperature conditions to predict equilibrium and reduce measurement time. These typically consist of exponential functions that are used to fit the measurement data. The fitted parameters then are used to predict the final equilibrium power.

8.1.1.3 The temperature of the item to be measured may be adjusted through the use of preconditioning heaters or baths in order to decrease the time required to reach equilibrium.

8.2 Heat-flow calorimeters are operated typically in one of two modes, passive mode or power compensation mode.

8.2.1 *Passive Mode Operation*—A plot with an example of a passive mode calorimeter response to a heat source is shown in Fig. 2. This plot shows that after a period of time the temperature transient caused by the insertion of the item into the calorimeter disappears and the calorimeter and item are in thermal equilibrium after approximately 7 h in this example.

8.2.1.1 The item power,  $W_{item}$ , is calculated by the following:

$$W_{item} = \left(BP_s - BP_0\right)/S \tag{2}$$

where:

- S = the calorimeter sensitivity ( $\mu$ V/W) at the power level, determined by electrical or <sup>238</sup>Pu standards,
- $BP_s$  = the equilibrium sensor response with the item in the calorimeter, and
- $BP_0$  = the baseline sensor response with no item in the calorimeter.

8.2.2 *Power Compensation Mode*—In this mode of operation, a constant amount of thermal power is applied to the item chamber by electrical heaters. The temperature of the calorimeter item chamber is held at a constant temperature difference above the temperature of the thermal sink by means

🖽 C1458 – 16



FIG. 2 Approach to Equilibrium for a Calorimeter in the Passive Mode

of a controlled power unit. The constant power causes a constant temperature differential to be maintained across the thermal resistance separating the measurement chamber from a controlled reference temperature. The temperature differential is proportional to the signal, voltage or resistance, and is the temperature difference between a sensor (or sensors) located adjacent to the item being measured and the other(s) located at the reference temperature. A closed-loop controller monitors the output signal, and if a radioactive heat-generating item is

inserted, the external power applied is decreased to precisely maintain the same signal differential. When the unknown item is placed in the calorimeter, the control power drops over time to a lower level. The power of the unknown is the difference between the two control power readings at equilibrium. A plot with an example of the calorimeter response is shown in Fig. 3.

8.2.2.1 The item power,  $W_{item}$ , is calculated by the following:



FIG. 3 Approach to Equilibrium for a Calorimeter Operated in the Power Compensation Mode

$$W_{item} = W_0 - W_H \tag{3}$$

where:

- $W_0$  = the base power with no item in the calorimeter, and
- $W_H$  = the power supplied to the calorimeter with the item in the calorimeter.

8.2.2.2 The measurement time for the power compensation mode of operation is normally shorter than for the passive mode because the calorimeter components are at the equilibrium temperature and the controlled internal heater can supply heat actively to bring the item to equilibrium.

### 9. Hazards

9.1 Safety Hazards:

9.1.1 It is recommended that a criticality evaluation be carried out if fissile material is to be measured.

9.1.2 Precautions should be taken to minimize electrical shock hazards.

9.1.3 Precautions should be taken to avoid contamination with radioactive materials.

9.1.4 Precautions should be taken to minimize personnel radiation exposure to ionizing radiation.

9.1.5 Pinch-point and lifting hazards may be present during the loading and unloading of heavy items with calorimeters. Mechanical aids, such as a hoist or any other assist, should be used for movement of heavy items.

9.1.6 High-power items could present a burn hazard to the operator or damages to the instrument.

#### 9.2 Technical Hazards:

9.2.1 Room temperature variations may affect the stability of the reference temperature and increase the measurement uncertainty.

9.2.2 Using a measurement result outside of the range of the calibration is not recommended.

9.2.3 Care should be taken in the insertion or removal of the calorimeter can so that it is not jammed in the calorimeter well.

9.2.4 Noise in the electronics AC supply power generated by nearby machinery may increase the measurement uncertainty.

9.2.5 The base power for power compensation mode calorimeters must not be less than the highest power expected from items.

9.2.6 Mechanical stress on the item chamber from the weight of the item may cause a bias in the final result.

9.2.7 The calorimeter may exhibit a small heat distribution error dependent on calorimeter design and item characteristics.

9.2.8 The following conditions could extend measurement time.

9.2.8.1 Large masses of material.

9.2.8.2 Items that make poor thermal contact with their containers.

9.2.8.3 Items that contain a large amount of insulating material or dead air spaces caused by several layers of containment.

9.2.9 Deviation between the actual approach to equilibrium and end-point prediction algorithm can result in measurement bias.

9.2.10 Bias in the determination of  $\mathrm{P}_{\mathrm{eff}}$  will lead to a bias in the assay result.

9.2.11 For removable electrical heat standards the heater leads to the standard will serve as a heat path for heat to exchange between the measurement chamber and the environment; this may bias the calibration.

9.2.12 Improperly closing the calorimeter (for example, not completely inserting the insulating baffle) may bias the measurement results, degrade precision, or both.

9.2.13 Cables or any other conductor providing a thermal conduction path into and out of the calorimeter may bias the measurement results. Variability in these paths is a more severe hazard.

## **10. Calibration Procedure**

10.1 The type of calibration procedure depends on whether the calorimeter is operated in the passive or power compensation mode. In the passive mode calibration consists of determining the calorimeter sensitivity, S, the conversion factor between the differential voltage or resistance output of the sensor system and the thermal power of the item being measured. In the power compensation mode calibration is setting the sensor output set point voltage that corresponds to a specific base power.

10.2 Calibration-Passive Mode:

10.2.1 Select a series of <sup>238</sup>Pu heat standards or calibrated electrical standard power settings that span the expected power range of items to be measured. A minimum of three different standard powers must be used.

10.2.2 Initiate a baseline measurement of the voltage with a calorimeter can, if used, filled with conductive material in the calorimeter chamber. There should be no heat source in the calorimeter can. Record the baseline,  $BP_0^{\rm pre}$ , after equilibrium is reached.

10.2.3 Remove the calorimeter can and place the appropriate standard in the measurement cavity as it is done for an unknown item.

10.2.3.1 Whether using radioactive or electrical heat standards, the calorimeter can must be removed from the calorimeter between each measurement, baseline, or standard. If possible, this removal is necessary even when using electrical standards to simulate as closely as possible real calorimeter operating conditions.

10.2.4 Close the calorimeter can and place it in the calorimeter well. For all measurements, the calorimeter should be properly closed before a measurement is commenced.

10.2.5 Initiate the calorimeter run. Record the voltage output  $(BP_{std})$  after equilibrium has been achieved.

10.2.6 Remove the calorimeter can from the calorimeter, then remove the heat standard from the can and place the can back in the calorimeter.

10.2.7 Re-measure the baseline  $BP_0^{\text{post}}$  after equilibrium is reached as needed.

10.2.8 Use the known power output of the heat standard,  $W_{std}$ , the calorimeter sensor value,  $BP_{std}$ , and the calculated average baseline (average of  $BP_0^{\text{pre}}$  and  $BP_0^{\text{post}}$ ),  $BP_0$  to calculate the calorimeter sensitivity, S.

$$S = \frac{BP_{std} - BP_{0(av)}}{W_{std}} \tag{4}$$

The sensitivity *S* may vary with the power of the standard,  $W_{std}$ , over the measurement range of the calorimeter. The change in sensitivity with power can be described by a mathematical function appropriate to the specific calorimeter response:

$$S = S_0 + k \times W_{std} \tag{5}$$

where:

- $S_o$  = the mathematically determined sensitivity for zero power, and
- k = the slope of the varying sensitivity.

As an example, the sensitivity for a nickel resistance wire-wound heat flow calorimeter of the Mound Gradient Bridge design is shown in Fig. 4.

10.2.9 Perform a minimum of three replicate measurements at each power level using steps 10.2.1 - 10.2.8. This sensitivity determination at different power levels typically need only be done once after a calorimeter is installed.

10.2.10 The equation  $W = f(\Delta V)$  relates sensor output to thermal power in Watts, where  $\Delta V$  is the measured voltage difference and f is an algorithm function chosen to the calorimeter design and historically has the form of a quadratic function (24). The relative standard deviation of an individual base power measurement should typically be less than 0.1 %.

#### 10.3 Power Compensation Mode:

10.3.1 Select an approximate base power, W, at which the servo system will operate. As a general rule, it should be 10 to 20 % higher than the highest expected item power.

10.3.2 Determine the sensitivity *S* of the calorimeter by performing steps 10.2.2 - 10.2.8 once for one <sup>238</sup>Pu or electrical standard power. Any power within the range of item powers is satisfactory.

10.3.3 The actual supplied heater power  $(W_0)$  used to maintain the set point bridge,  $BP_{sp}$  may be slightly different than the target value, W, taking into account the uncertainty in S.

10.3.4 Place a calorimeter can, if used, that is filled with conductive material but with no heat generating item in the calorimeter.

10.3.5 Close the calorimeter and initiate the measurement run. Once equilibrium is reached take a heater power measurement. This is the base power  $W_0$  corresponding to  $BP_{sp}$ .

10.3.6 Repeat the base power measurement with no item in the item chamber at least three times. This involves removing and reinserting the calorimeter can between each of the base power measurements. The relative standard deviation of an individual base power measurement should typically be less than 0.1 %.

## 11. Assay Procedure

11.1 Item Measurement—Determine the baseline voltage  $BP_0$  or base power  $W_0$  for the calorimeter. This should be done using step 10.2.2 for the passive mode or steps 10.3.4 and 10.3.5 for the power compensation mode prior to the item measurement. If  $BP_0$  or  $W_0$  is stable the baseline or base power measurement need not be done prior to every item measurement. Time between zero power measurements are set based on calorimeter stability, required accuracy, and administrative requirements.

11.1.1 Load the item to be assayed into the calorimeter can, and close the can.

11.1.2 Load the can into the calorimeter measurement chamber.

11.1.3 Properly close the calorimeter.

11.1.4 Initiate the measurement.

11.1.5 Terminate the measurement when the thermal equilibrium has been established or predicted.

11.1.6 An additional baseline (passive) or base power (power compensation) run may optionally be taken after the item measurement-using step 10.2.2 or steps 10.3.4 and 10.3.5. The average of the pre- and post-item measurement baselines or base powers may be used for  $BP_0$  or average base power  $W_0$ .



The grouping of points at 1 W are from the CALEX PuO<sub>2</sub> standard; the rest of the points are from <sup>238</sup>Pu heat sources (23). **FIG. 4 The Differential Sensitivity of the ARIES I Calorimeter Shows a Sensitivity of About 20 000 μV/W** 

(L) C1458 – 16

11.2 *Item Thermal Power Calculation*—The calculation of the item thermal power is performed differently for passive mode compared to power compensation mode.

11.2.1 *Power Calculation–Passive Mode*—Calculate the item power using the equation for  $W_{item}$  (24).

$$W_{item=\frac{(BP_{item} - BP_0)}{S}}$$
(6)

11.2.2 Power Calculation–Power Compensation Mode:

11.2.2.1 Calculate the item power,  $W_{item}$  as follows:

$$W_{item} = W_O - W_H \tag{7}$$

where:

 $W_O$  = control power with no item in the calorimeter, and  $W_H$  = control power with an item in the calorimeter.

11.2.3 If necessary, a bias correction may be made to a calorimeter measurement based on standard measurements made under a measurement control program. It is unusual if a calorimeter requires a bias correction. Proper operation of the calorimeter hardware and software should be definitively verified before resorting to bias corrections.

11.2.4 Once the thermal power is determined the quantity of radioactive material in the container is calculated using the equations in 11.3.1. For Pu, an isotopic analysis is required to determine  $P_{eff}$ .

#### 11.3 Calorimetric Assay—Plutonium:

11.3.1 The mass of plutonium in an item can be calculated from the thermal power measured by calorimetry in units of W and from the item effective specific power in units of W/g Pu as shown in Eq 8:

$$m = \frac{W_{item}}{P_{eff}} \tag{8}$$

 $P_{eff}$  is the important factor required to convert the measured thermal power in Watts from the calorimeter to grams of elemental plutonium.  $P_{eff}$  is most often determined from an independent knowledge of the plutonium isotopic composition and <sup>241</sup>Am content of the measured item. The isotopic composition information can come from destructive mass spectrometry measurements, or nondestructively from gamma ray isotopic analysis measurements.

11.3.2 The effective specific power (W/g Pu) of the material in the item is calculated in Eq 9:

$$P_{eff} = \sum_{i} R_{i} \times P_{i} \tag{9}$$

where:

- $R_i$  = abundance of the *i*-th isotope (i = <sup>238</sup>Pu, <sup>239</sup>Pu, <sup>240</sup>Pu, <sup>241</sup>Pu, <sup>242</sup>Pu, and <sup>241</sup>Am) of the Pu in the item expressed as a weight fraction, g  $m_i/m_{Pu}$ , and
- $P_i$  = a physical constant, the specific power of the *i*-th isotope in the item in W/g.

<sup>241</sup>Am mass is determined using the Am:Pu ratio:

$$M_{Am} = R_{Am} \cdot M_{Pu} \tag{10}$$

where:

 $R_{Am}$  = the mass ratio of <sup>241</sup>Am to Pu, and  $M_{Pu}$  = the mass of Pu.

11.3.2.1 The isotopic fractions,  $R_i$ , can be determined destructively by mass spectrometry and alpha spectrometry, or nondestructively by gamma-ray spectrometry. Liquid scintillation counting may be used as well as alpha spectrometry for determination of the <sup>241</sup>Am abundance by means of destructive analysis. The terms  $R_iP$  are the contribution of each isotope to the total item power. The isotopic fractions  $R_i$  are not constant, but change as the constituent isotopes in the item decay or grow in.

11.3.2.2 The isotopic specific powers,  $P_i$ , for relevant Pu isotopes have been directly measured or can be derived from Eq 11. The specific powers and half-lives for Pu isotopes and <sup>241</sup>Am are listed in Table 1. The half-lives were established after a data review of collaborative and individual experiments by the USDOE Half-Life Evaluation Committee. The half-lives and specific powers of <sup>239</sup>Pu and <sup>240</sup>Pu were determined from collaborative experiments on nearly pure isotopic samples. The specific powers of <sup>241</sup>Pu and <sup>242</sup>Pu were measured by individual experimenters and the specific power of <sup>238</sup>Pu and <sup>241</sup>Am were determined using Eq 11:

$$P_{i} = \frac{2119.3}{T_{1/2}^{i}A_{i}} \times Q_{i}$$
(11)

where:

 $P_i$  = specific power of the isotope *i* (W/g),

 $Q_i$  = total disintegration energy (MeV) for alpha-particle emitters, or the average energy (MeV) of beta particle emitters of pure beta emitter,

 $T_{i^{1/2}}$  = half-life (years) of isotope *i*, and

		-		-	
Isotope	Half-Life, Years	RSD (%)	Specific Power (W/g)	RSD (%)	References
<sup>238</sup> Pu	87.74	0.05	0.56757	0.05	(25, 26)
<sup>239</sup> Pu	24 119	0.11	1.9288 × 10 <sup>-3</sup>	0.02	(26-28)
<sup>240</sup> Pu	6564	0.17	7.0824 × 10 <sup>-3</sup>	0.03	(29-34)
<sup>241</sup> Pu	14.348	0.15	3.412 × 10 <sup>−3</sup>	0.06	(35-39)
<sup>242</sup> Pu	376 300	0.24	0.1159 × 10 <sup>−3</sup>	0.22	(40)
<sup>241</sup> Am	433.6	0.32	0.1142	0.37	(38, 41)

TABLE 1 Nuclear Decay Parameters for Pu Calorimetric Assav<sup>A</sup>

<sup>A</sup>ANSI N15.22, the original compiled source for these parameters, is no longer an active ANSI standard. The References are those from ANSI N15.22 and are the appropriate references for the values in Table 1.

 $A_i$  = molar mass of nuclide *i* (g/mol).

11.3.2.3 A recent publication (42) of a reanalysis of  $^{241}$ Pu half-life data gives a value of 14.329 ± 0.015 (SD), in agreement with the value in Table 1. While other references give slightly different values for some of the half-lives in Table 1, an analysis of the effects of these half-life variations (43) shows very little effect on the overall sample power giving support for continuing to recommend use of the values in Table 1.

## 12. Measurement Control Procedure

12.1 Measurement control procedures are similar for all calorimeter measurements.

12.1.1 Replicate measurements of <sup>238</sup>Pu or electrical heat standards are performed to demonstrate that the calorimeter system, hardware and software, is operating correctly. Other well-characterized nuclear materials may be used as heat standards.

12.1.2 Corrections to the thermal power of radioactive heat standards due to radioactive decay must be made.

12.1.3 Control charts monitoring from replicate measurements of heat standards can be beneficial in providing quantitative means for determining that the calorimeter system is operating satisfactorily prior to a single assay measurement or group of assay measurements. In addition these charts can be used to demonstrate that the calorimeter was in control during the assay measurements.

12.1.3.1 Control charts monitoring may also be used to monitor baseline or base power measurements. This can provide auxiliary information in case out-of-control operating conditions are detected.

12.1.4 Calorimeter bath temperatures can be monitored continuously to flag changes that will affect calorimeter performance. The temperatures may be evaluated using control charts or administrative limits.

12.1.5 The frequency of standards and baseline measurements are dependent on how well the calorimeter meets performance requirements and environmental conditions. Other factors to consider are throughput/day, portable or permanent instruments, and data collection systems.

12.1.5.1 More details applicable to calorimeter measurement control may be found in ANSI N15.36–2010. More general aspects of measurement control relevant to calorimetry may be found in Guide C1009.

12.1.6 Data collected from a measurement control program can be used to calculate the precision and bias of the power measurement. As an example, a summary of the precision and bias of the power measurement obtained from replicate measurements of <sup>238</sup>Pu heat standards in production facilities over a 0.5–1.0-year period is shown in Table 2. Extensive calorimetric assay precision and bias data can be found in references summarized in (**15, 44**).

#### 13. Assay Precision and Bias

13.1 Precision:

TABLE 2 Calorimeter Power Measurement Precision and Bias (15, 45, 46)

Heat Standard Power, W	Calorimeter Diameter, m	Calorimeter Type, Operation Mode	Number of Measurements	Precision, % RSD	Bias, %
98	0.06	rod, servo (15)	29	0.065	0.02
3.5	0.15	rod, servo (15)	55	0.09	0.00
4.0	0.25	twin, passive <sup>A</sup> (15)	22	0.05	0.03
4.9	0.30	twin, passive <sup>A</sup> (15)	34	0.06	0.05
0.0786	0.04	Solid state, passive <sup>B</sup>	10	0.23	0.001
0.01	0.63	Mono cell passive (45, 46)		12	15
0.05	0.63	Mono cell passive (45, 46)		1.8	2.5
0.1	0.63	Mono cell passive (45, 46)		0.9	2
0.3	0.63	Mono cell passive (45, 46)			2

<sup>A</sup> Pooled results from two calorimeters.

<sup>B</sup>Measurements made in laboratory.

13.1.1 Calorimetric assay of plutonium requires both calorimetry and isotopic analysis, thus, the precision and bias of the assay will have components due to uncertainties in both techniques.

13.1.2 The precision of a calorimeter measurement is dependent on the sensitivity of the calorimeter, baseline stability, and item power.

13.1.3 A list of major factors that can affect the precision of the gamma-ray isotopic assay and mass spectrometric modes of isotopic analysis are described in Test Methods C1030 and C697, respectively.

13.1.4 Plutonium in  $PuO_2$ - $UO_2$  Mixed Oxide—Generally, the greater the thermal power in a calorimeter, the better the relative precision. Absolute calorimeter random noise increases with increasing power. To illustrate this the relative precision observed from repetitive calorimeter measurements of six items containing 26 to 258 g of Pu (17 % <sup>240</sup>Pu) in PuO<sub>2</sub>-UO<sub>2</sub> (26 % Pu) was calculated and the results are shown in Table 3. These measurements were made over a 56-day period with a water bath twin-bridge over-under calorimeter. The items were

 
 TABLE 3 Calorimetry/Mass Spectrometry Measurements of Plutonium in PuO<sub>2</sub>-UO<sub>2</sub> Mixed Oxide<sup>A,B</sup> (15, 47)

ID	Mass <sup><i>B,C</i></sup> , g by Chem	Mass, g <sup>C</sup> , by Cal <sup>D</sup> /MS/α <sup>E</sup>	Precision, g 1 SD	Precision %RSD <sup>F</sup>	Bias, %
4	257.70	257.54	0.14	0.06	-0.06
5	206.09	206.06	0.13	0.06	-0.02
6	206.18	206.12	0.14	0.07	-0.03
7	128.81	128.94	0.12	0.09	0.10
8	77.28	77.35	0.12	0.15	0.09
9	25.79	25.99	0.11	0.42	0.78

<sup>A</sup> Unpublished results.

<sup>B</sup> Mass of plutonium determined by coulometry using reference material NBS 949E. Pu fraction of mixed oxide, 0.25759, based on triplicate measurements of six samples.

<sup>C</sup> Pu masses reported here decayed to a common date.

<sup>D</sup> Final results based on 117 replicate calorimeter measurements/item.

 $^{\it E}$  Isotopic composition determined by 12 replicate measurements by mass spectrometry ( $^{239}$ Pu,  $^{240}$ Pu,  $^{241}$ Pu,  $^{242}$ Pu) and six replicate alpha counting ( $^{238}$ Pu,  $^{241}$ Am) measurements. Average Pu isotopic and  $^{241}$ Am results were used to calculate  $P_{\rm eff}$  used for all calorimeter measurements.

<sup>F</sup> Precision due to calorimetry power replicate measurements.

loaded robotically allowing for continuous operation. A common  $P_{eff}$  factor for all six items was determined using the plutonium isotopic composition and <sup>241</sup>Am content that was determined by mass spectrometry and alpha counting. The calorimeter can size was 0.06-m diameter  $\times$  0.16-m high. The calorimeter measurement time was fixed at 1 h. The calorimeter was used in the power compensation mode and the items were preconditioned to reduce measurement time.

13.1.5 Plutonium in PuO2-The results of multiple calorimetric and gamma-ray isotopic measurements by three facilities on identical standards each containing 400 g of wellcharacterized Pu (6 % <sup>240</sup>Pu) in PuO<sub>2</sub> are shown in Table 4. The Pu content and isotopic composition used as reference values of the mother lot of PuO<sub>2</sub> material used for these standards were measured by coulometry and mass spectrometry/alpha counting by four analytical laboratories. The calorimetry and gamma-ray measurements used to determine Pu mass variabilities and biases reported in Table 4 were taken over a one-year period. The within-facility variability and the bias of the calorimetric/gamma-ray assay were calculated from results reported by each facility decayed to a common date. Each facility used different gamma-ray analysis codes for the isotopic measurements. Some of the reported values are the averages of measurements of the standard item with different calorimeters.

#### 13.2 Bias:

13.2.1 The bias of calorimetric assay can be determined by the measurement of certified reference materials or wellcharacterized items with known elemental and isotopic compositions.

13.2.2 Plutonium in  $PuO_2$ - $UO_2$  and  $PuO_2$  (44, 48). Biases for calorimetric assay of Pu in  $PuO_2$ - $UO_2$  mixed-oxide powder and Pu in  $PuO_2$  powder are shown in Tables 3 and 4, respectively. This data shows that the biases for this measurement application are typically less than 0.1 %.

13.2.3 The algorithm used for equilibrium prediction that is suitable for a <sup>238</sup>Pu heat standard or electrical standard in a conductive matrix may not be suitable for items with Pu in a matrix with poor thermal conductivity.

 TABLE 4 Calorimetry/Gamma-Ray Assay Measurement of 400 g

 of Pu in PuO2<sup>A</sup> (15, 20, 48)

Facility	Within-Facility Variability, g	Within-Facility Variability, %RSD	Bias, g	Bias, %
A <sup>B</sup>	1.5	0.38	0.03	0.01
B <sup>B</sup>	1.5	0.38	-0.40	-0.10
$C^{c}$	1.4	0.36	0.04	0.01

<sup>A</sup> All masses are in grams of plutonium decayed to a common date.

<sup>B</sup> Used multiple water bath twin-bridge calorimeters.

<sup>C</sup> Used "air-bath" calorimeter.

13.2.4 *Plutonium in Salt Residues* (16, 49, 50)—In two separate studies calorimetry and gamma-ray spectroscopy measurements were used to assay items containing a mixture of KCl, MgCl<sub>2</sub>, and NaCl residues containing PuCl<sub>3</sub> and AmCl<sub>3</sub>, as well as, shards of MgO crucibles and plutonium metal shot. The material in each item was sampled and analyzed for Pu. The average relative biases between the calorimetry assay and alternative assay is shown in Table 5. A twin bridge water bath calorimeter was used for the thermal power measurements in both studies.

13.2.5 Americium in Salt Residues (16, 49, 50)—The same measurement techniques described in section 13.2.4 were used to assay the <sup>241</sup>Am content of the items containing molten salt residues, as well as, the Pu content. Gamma-ray spectroscopy was used to determine the <sup>241</sup>Am/Pu ratio simultaneously with plutonium isotopic ratios. The biases are shown in Table 5.

13.2.6 Tritium (18, 51)—Calorimetry was used to measure the quantity of tritium gas in containers. Since tritium was the only radioactive isotope, no isotopic measurements were required for the assays. After the calorimeter measurement the gas was quantitatively transferred to tanks with calibrated volumes, and the quantity of tritium was determined using calibrated pressure and temperature transducers and mass spectrometric analyses. A comparison of measurement results between calorimetry and pressure volume temperature (PVT) combined with mass spectrometry (MS) was made for 50 containers. The tritium content of the containers ranged from 15 to 16 g. The relative mean bias for the calorimeter assay compared to PVT combined with MS was –0.12  $\pm$  0.05 % average relative standard deviation ( $RSD_{mean}$ ). A twin bridge water bath calorimeter was used for the calorimeter measurements.

#### 14. Keywords

14.1 accountability; calorimetry; inventory; nondestructive assay; plutonium; quantification; tritium

TABLE 5 Bias—Calorimetry/Isotopic Measurements of Pu and Am in Molten Salt Residues

Mass Range, g Pu	No. of Items	Mass Range, g Am	Bias, % Pu	Bias, % Am
35-416	10	0.4–18	1.5 ± 0.5 <sup>A</sup>	$0.2 \pm 0.4^{A}$
214-414	9	5–26	$1.6 \pm 0.6^{B}$	$0.2 \pm 0.6^{B}$

<sup>A</sup>Chemistry—Pu and Am elemental analysis by Isotopic Dilution Mass Spectrometry. Aliquots taken of blended salt matrix for analysis. Pieces of Pu metal removed, oxidized, and returned to matrix before blending.

Calorimetry—Isotopic analysis for  $P_{\rm eff}$  determination by gamma-ray spectroscopy.

<sup>B</sup> Chemistry—Entire matrix of each item dissolved and liquid samples measured by x-ray fluorescence for Pu and gamma counting for Am.

Calorimetry—Isotopic analysis for  $P_{\rm eff}$  determination by gamma-ray spectroscopy and facility stream averages of certain Pu isotopic ratios.

#### APPENDIXES

#### (Nonmandatory Information)

## X1. UNCERTAINTY ESTIMATION FOR PLUTONIUM AND <sup>241</sup>AM MASS

X1.1 In Eq 8, the measurement of the thermal power, W, is independent of the measurement of the effective specific power,  $P_{eff}$ , and thus, the relative uncertainty for the plutonium mass, M, can be written as follows:

$$\frac{\sigma_M}{M} = \left[\frac{\sigma_W^2}{W^2} + \frac{\sigma_{P_{eff}}^2}{P_{eff}^2}\right]^{1/2}$$
(X1.1)

X1.1.1 The uncertainty in the power measurement,  $\sigma_W$ , can be obtained from replicate power measurements of heat standards or from historical data. It should include both precision and bias components. The uncertainty in  $P_{eff}$ ,  $\sigma_{Peff}$ , comes from the uncertainty in the isotopic fractions,  $R_i$ , and isotopic specific powers,  $P_i$ . The uncertainties in the isotopic fractions are determined from uncertainties in the various techniques that might be used for the isotopic analysis, such as mass spectroscopy, alpha counting, or gamma-ray spectroscopy. There are sufficient gamma-rays in plutonium to provide independent measured isotopic ratios of the major contributors to the item thermal power:  ${}^{238}$ Pu,  ${}^{240}$ Pu,  ${}^{241}$ Pu,  ${}^{242}$ Pu, and  ${}^{241}$ Am with respect to  ${}^{239}$ Pu that allow  $R_i$  to be calculated. Ref (52) discusses this in more detail. The uncertainties in the isotopic specific powers,  $P_i$ , as determined by different experiments, are given in Table 1. The test method for determining isotopic composition by gamma-ray spectroscopy is described in Test Method C1030. Several gamma-ray analysis codes are available commercially that not only provide the isotopic composition, but also the uncertainties of the isotopic fractions and the specific power of the item being measured. Error propagation of the isotopic fractions is discussed in Ref (52) although the discussion in Ref (14) applies strictly only to the case where the measured isotopic ratios are completely independent.

X1.1.2 The uncertainty of the <sup>241</sup>Am mass mixed with Pu is as follows (53):

$$\frac{\sigma_{M}}{M} = \left\{ \left[ \left( \frac{\sigma_{W}}{W} \right)^{2} + \left( \frac{K}{P_{eff}} \right)^{2} \left[ \left( \frac{\sigma_{K}}{K} \right)^{2} + \left( \frac{\sigma_{R_{Am}}}{R_{Am}} \right)^{2} \right] + \left( \frac{R_{Am} \sigma_{P_{Am}}}{P_{eff}} \right)^{2} \right] \right\}^{1/2}$$
(X1.2)

where:

$$K = P_{\rm eff} - P_{\rm Am} R_{\rm Am} \tag{X1.3}$$

where:

 $P_{Am}$  = the specific power of <sup>241</sup>Am,  $R_{Am}$  = the mass ratio of <sup>241</sup>Am to plutonium, W = the thermal power

= the thermal power,

 $\sigma_W$  = the uncertainty in the thermal power,

 $\sigma_{R_{AM}}$  = the uncertainty in the <sup>241</sup>Am mass ratio, and  $\sigma_{P_{Am}}$  = the uncertainty in the <sup>241</sup>Am specific power.

 $\sigma_{P_{Am}}$ 

$$\sigma_{K} = \left(\sum_{i \neq A_{m}} (\sigma_{P_{i}}^{2} \cdot R_{i}^{2} + \sigma_{R_{i}}^{2} \cdot P_{i}^{2})^{1/2} \right)$$
(X1.4)

Where  $\sigma_{P_i}$  comes from Table 1 and the  $\sigma_{R_i}$  are taken from the isotopic code report.

## **X2. UNCERTAINTY ESTIMATION FOR TRITIUM MASS**

X2.1 The uncertainty of a calorimetric assay of tritium can be calculated using Eq X1.1. The uncertainty of the effective specific power,  $P_{\rm eff}$ , of tritium is the same as the isotopic specific power, 0.00045 W obtained from 4.4.1. Dividing by the specific power of tritium, 0.3240 W/g, results in the following:

$$\frac{\sigma_{P_{\text{eff}}}}{P_{\text{eff}}} = 0.0014$$
 (X2.1)

So for tritium the relative uncertainty of the tritium mass is as follows:

$$\frac{\sigma_{\text{tritium}}}{M_{\text{tritium}}} = \left[ \left( \frac{\sigma_W}{W} \right)^2 + (0.0014)^2 \right]^{1/2}$$
(X2.2)

6 C1458 – 16

#### REFERENCES

- (1) Pillinger, W. L., Hentges, J. J., and Blair, J. A., "Tritium Decay Energy," *Phys. Rev.*, 1961, pp. 232–233.
- (2) "Handbook of Nuclear Safeguards Measurements," Ed. D. R. Rogers, NUREG/CR-2078, 1981, pp. 533–550.
- (3) "Passive Non-Destructive Assay of Nuclear Materials," Ed., Doug Reilly, Norbert Ensslin, and Hastings Smith, Jr., NUREG/CR-5550, 1991, pp. 641–660.
- (4) Mason, J. A., "The Use of Calorimetry for Plutonium Assay," UKAEA Report SRDP-R100, 1982.
- (5) "In-Field Calibration of Neutron Correlation Counter via Calorimetry and High Count Rate Gamma-Ray Isotopic Abundance Measurements," IAEA-SM-293/126, *Nuclear Safeguards Technology*, 1986, Vol 2, IAEA, 1987, pp. 239–249.
- (6) "In Field Calibration of a Neutron Correlation Counter Via Calorimetric Assay," 2<sup>nd</sup> International Workshop on Calorimetric Assay, 27 October 1994, Santa Fe, New Mexico, USA.
- (7) Eichelberger, J. F., Jordan, K. C., Orr, S. R., and Parks, J. R., "Calorimetric Determination of the Half-Life of Polonium-210," *Physical Review*, 96, 719, 1954.
- (8) Gunn, S. R., "Radiometric Calorimetry: A Review," Nuclear Instruments and Methods 29, 1964, 1.
- (9) Gunn, S. R., "Radiometric Calorimetry: A Review" (1970 supplement), Nuclear Instruments and Methods 85, 1970, 285, 1970.
- (10) Cremers, T. and Wachter, J. R., "Selection of Nondestructive Assay Methods: Neutron Coincidence or Calorimetric Assay?" *Nuclear Materials Management*, Vol XIII, pp. 483–486, Proceedings Issue, 1994.
- (11) Lestone, J. P., Prettyman, T. H., and Chavez, J. D., "Performance of the Skid-Mounted Tomographic Gamma Scanner for Assays of Plutonium Residues at RFETS," Proceedings of the Institute of Nuclear Material Management 41st Annual Meeting, July 16-20, 2000, New Orleans, Louisiana, USA.
- (12) Hyman, P., Gupta, V., and Sullivan, D., "Nuclear Materials Safeguards Matrix-Specific Qualification & Continuous Bias Correction Programs, Minutes of the Measurement Evaluation Program Meeting," New Brunswick Laboratory, Argonne, IL, May 26-27, 1999.
- (13) Perry, R. B., Fiarman, S., Jung, E. A., and Cremers, T., "Test and Evaluation of the Argonne BPAC10 Series Air Chamber Calorimeter Designed for 20 Minute Measurements," ANL-90/36, 1990.
- (14) Jossens, G., Mathonat, C., Bachelet, F., "Waste management nuclear calorimeter for very large sample," INMM 2013, 54<sup>th</sup> Annual Meeting, July 14-19, 2013, Palm Spring, California, USA.
- (15) Bracken, D.S., and Rudy, C. R., Principles and Application of Calorimetry Assay, PANDA 2007 Addendeum http://www.lanl.gov/ orgs/n/n1/panda/.
- (16) Bracken, D. S., Biddle, R. S., Carrillo, L. A., Hypes, P. A., Rudy, C. R., Schneider, C. M., Smith, M. K., "Application Guide to Safeguards Calorimetry," LA-13867-M, January (2002).
- (17) Thornton, M. I. and Schillebeeckx, P., "Performance Evaluation of a Thermopile Small Sample calorimeter in PERLA," ESARDA, 23<sup>rd</sup> annual meeting, 2001 May 8-10, Bruges, Belgium.
- (18) Croft, S., Philips, S., Evans, L. G., Guerault, J., Jossens, G., Passelegue, L., Mathonat, C., "Nuclear Calorimetry – Dispelling the Myths," WM2010, March 7-11, 2010, Phoenix, Arizona, USA.
- (19) Carrillo, L., Rudy, C. R., Long, S. M., McDaniel, J., and Rodenburg, W. W., "Heat Standards Calibration in the DOE Complex," *Nuclear Materials Management*, Vol XXVI Proceedings Issue, 1997.
- (20) Renz, D. P., Wetzel, J. R., Severinghaus, M. B., Sheldon, R. I., and Bruchner, L., "High Wattage Comparison of the MECH Probe and <sup>238</sup>Pu Heat Standards at Los Alamos," *Nuclear Materials Management*, Vol XXIII, Proceedings Issue, 1994, pp. 95.
- (21) Mason, J. A. and Thornton, M. I., "Electrical and Heat Standard Calibration of Calorimeters for Plutonium and Tritium

Measurement," Proceedings of the 17<sup>th</sup> ESARDA Symposium on Safeguards and Nuclear Materials Management, Aachen, May 1995, pp. 581–583.

- (22) Mason, J. A., Fleissner, J. G., Valdez, J. L., Campbell, G., and West, J., "High Precision Large Sample Plutonium Calorimetry Measurement," *Nuclear Materials Management*, Vol XXVI, Proceedings Issue, 1997.
- (23) Sampson, T. E., Teresa, L., Cremers, T., Wenz, T. R., Hansen, W. J., Scheer, N. L., Kelley, T. A., Harker, W. C., Mayo, D. R., Bracken, D. S., Herrera, G. D., and Guillen, B. A., "The ARIES Nondestructive Assay System: Description and Process Demonstration Results," Los Alamos National Laboratory report LA-14143, May 2004.
- (24) Croft, S., Hauck, D. K. and Bracken, D., "On the Functional Form used to Represent the Response of Nuclear Calorimeters," ESARDA-2011, May 15, 2011, Budapest, Hungary.
- (25) Wapstra, A. H. and Bos, K., "The 1977 Atomic Mass Evaluation," *Atomic Data and Nuclear Data Tables*, Vol 19, No. 3, 1977.
- (26) Strohm, W. W., "The Measurement of the Half-Life of Plutonium-239 by the U. S. Half-Life Evaluation Committee," *International Journal of Applied Radiation and Isotopes*, Vol 29, 1978, pp. 481–483.
- (27) Seabaugh, P. W. and Jordan, K. C., "Calorimetric Determination of the Half-Life of Plutonium-239," *International Journal of Applied Radiation and Isotopes*, Vol 29, 1978, pp. 489–496.
- (28) Gunn, S. R.. "A Calorimetric Determination of the Specific Power of Plutonium-239," *International Journal of Applied Radiation and Isotopes*, Vol 29, 1978, pp. 497–499.
- (29) Rudy, C. R., Jordan, K. C., and Tsugawa, R., "Calorimetric Determination of the Half-Life of Plutonium-240," *International Journal* of Applied Radiation and Isotopes, Vol 35, 1984, pp. 177–180.
- (30) Lucas, L. L., and Noyce, J. R., "The Half-Life of Plutonium-240," *International Journal of Applied Radiation and Isotopes*, Vol 35, 1984, pp. 173–176.
- (31) Beckman, R. J., Marsh, S. F. Abernathy, R. M., and Rein, J. E., "Plutonium-240 Half-Life Determined by Isotope-Dilution Mass Spectrometric Measurement of the Grown-in Uranium-236," *International Journal of Applied Radiation and Isotopes*, Vol 35, 1984, pp. 163–169.
- (32) Steinkruger, F. J. Matlack, G. M., Beckman, R.T., "The Half-Life of <sup>240</sup>Pu Determined by Specific-Activity Measurements," *International Journal of Applied Radiation and Isotopes*, Vol 35, 1984, pp. 171–172.
- (33) Jaffey, A. H., Diamond, H., Bentley, W. C., Graczyk, D. G., Flynn, K. F., "Half-Life of <sup>240</sup>Pu," *Physical Review*, C, Vol 18, 1978, pp. 969–974.
- (34) Strohm, W. W., "The Measurement of the Half-Life of Plutonium-240 by the U. S. Half-Life Evaluation Committee," *International Journal of Applied Radiation and Isotopes*, Vol 35, 1984, pp. 155–157.
- (35) Marsh, S. F., et al, "Plutonium-241 Half-Life Determined by Mass Spectrometric Measurements of <sup>241</sup>Pu: <sup>242</sup>Pu Ratios," *International Journal of Applied Radiation and Isotopes*, Vol 31, 1980, pp. 629–631.
- (36) Garner, E. L., Machlan, L. A., "Mass Spectrometric Measurements to Determine the Half-Life of <sup>241</sup>Pu," NBS Special Publication 582, National Bureau of Standards, 1980, June, pp. 34–40.
- (37) DeBievre, P., "Isotope Dilution Mass Spectrometry; Some Present Limitations and Possibilities to Overcome These," Minutes of SALE program participants meeting, 1981 July 8–9 pp. 188–207.
- (38) Jordan, K. C., unpublished results, Monsanto Research Corporation, Mound Laboratory, Miamisburg, OH (1982).
- (39) Oetting, F. L., "Average Beta Energy of Plutonium-241 by Calorimetry," *Physical Review*, Vol 168, No. 4, 1968, pp. 1398–1401.

- (40) Osborne, D. W., Flotnow, H. E., "Half-Life of <sup>242</sup>Pu from Precise Low Temperature Heat Capacity Measurements," *Physical Review C*, Vol 14, No. 3, 1976, pp. 1174–1178.
- (41) Oetting, F. L., Gunn, S. R., "A Calorimetric Determination of the Specific Power and half-life of <sup>241</sup>Am," *Journal of Inorganic and Nuclear Chemistry*, Vol 29, 1967, pp. 2659–2664.
- (42) Croft, S., Burr, T. L., Favalli, A., "Estimating the Half-Life of <sup>241</sup>Pu and its Uncertainty," *Radiation Measurements* 59 (2013), pp. 94-102.
- (43) Srinivasan, B., Soriano, M., and Losinger, W. "Effects of Variations in Half-Lives on Decay-Corrected Characterized Values of Plutonium Standards in Calorimetric Exchange Program," Proceedings of 47th INMM, 2006.
- (44) Srinivasan, B., Waggoner, J., Mathew, K., Buerger, S., Narayanan, U. and Neuhoff, J., Calorimetry Exchange Program, Annual Report, NBL-ME 2008-Calex, July 2008.
- (45) Bachelet F., Dujardin, L., Clouard, S., Collardey, A., Theobald, M., Jossens, G., Mathonat, C., Etherington, G., "Large Nuclear Calorimeter for Non Destructive Assay on 55 Gallons Tritiated Waste Drums," INMM 55<sup>th</sup> Annual Meeting, July 20-24, 2014, Atlanta, Georgia, USA.
- (46) Jossens, G., Mathonat, C., Bachelet, F., "Nuclear Waste Calorimeter for Very Large Drums with 385 Litres Sample Volume," *Fusion Science and Technology*, Vol. 67, No 2, March 2015, 390-393.
- (47) Briden, J., Duff, M. F., Fultz, R. L., Kesling, W. E., Wetzel, J. R.,

Campbell, A. R., Fellers, C., Lemming, J. F., and Rudy, C. R., "Automated Plutonium Assay System (APAS) Redux," Proceedings of the Institute of Nuclear Material Management, 41<sup>st</sup> Annual Meeting, July 16-20, 2000, New Orleans, Louisiana, USA.

- (48) "Calorimetry Exchange Program, Quarterly/Annual Data Report, Calendar Year 1998," prepared by M. Irene Spaletto and David T. Baran, NBL-353, May 1999.
- (49) Longmire, V. L., Cremers, T. L., Sedlacek, W. A. Long, S. M., Scarborough, A. M., and Hurd, J. R., "Isotopic Ratios and Effective Power Determined by Gamma-Ray Spectroscopy vs. Mass Spectroscopy for Molten Salt Extraction Residues," *Nuclear Materials Management*, Vol. 31, Proceedings Issue, 1990.
- (50) Fleissner, J. G. and Hume, M. W., "Comparison of Destructive and Nondestructive Assay of Heterogeneous Salt Residues," RFP-3876, 1986.
- (51) Lindsay, C. N. Sprague, R. E., and Brandenburg, J. A, "A Measurement Control Study for Tritium Gas," MLM-2441, 1987.
- (52) Sampson, T. E., Gunnink, R. "The Propagation of Errors in the Measurement of Plutonium Isotopic Composition by Gamma-Ray Spectroscopy," *Journal of the Institute of Nuclear Materials Management*, Vol XII, No 2, 1983 Summer, pp. 39–48.
- (53) Croft, S., Nicholson, A., Kirkpatrick, J. M., Sampson, T. E., and Burr, T. "The Influence of Correlated Variance Structure in Relative Isotopic Abundances on Nondestructive Assay Results," INMM 56<sup>th</sup> Annual Meeting, July 12-16, 2015, Indian Wells, California, USA.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/