



Standard Specification for Commercially Packaged Laboratory Apparatus¹

This standard is issued under the fixed designation E920; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

This specification is proposed as a package performance standard to be met for items shipped domestically. The standard would apply only when requested by the buyer.

1. Scope

1.1 This specification covers the procedures for testing loaded shipping containers. Drop and vibration tests are performed to measure the ability of the shipping container to protect the product from the shock and vibration the container receives during normal handling and transporting. This specification is not intended to supplant material specifications or existing preshipment test procedures. *This specification is not intended for use with hazardous materials.*

1.2 These procedures are suitable for all types of laboratory apparatus, including reusable and disposable macro and micro products.

1.3 The values stated in inch-pound units are to be regarded as standard. The values given in parentheses are mathematical conversions to SI units that are provided for information only and are not considered standard.

1.4 The following precautionary caveat pertains only to the test method portion, Section 4, of this specification: *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards*:²

[D685 Practice for Conditioning Paper and Paper Products for Testing](#)

[D999 Test Methods for Vibration Testing of Shipping Containers](#)

[D3951 Practice for Commercial Packaging](#)

[D5276 Test Method for Drop Test of Loaded Containers by Free Fall](#)

3. Requirements

3.1 Three test specimens are required. Each specimen shall be run through the sequence of tests in the order given.

3.2 Condition test specimens in accordance with Practice [D685](#). Standard conditions must be maintained throughout the test sequence.

3.3 The packaging shall comply with Practice [D3951](#) except cleanliness of laboratory apparatus shall be as required by the Product Standard, the Quality Assurance Standard, or as agreed upon between the manufacturer and the purchaser.

4. Procedure

4.1 *Vibration Test*—Perform the vibration test in accordance with Method A of Methods [D999](#).

4.1.1 Proceed with the following test duration and sequence: bottom for 30 min; side for 15 min; and end for 15 min. Total vibration time will be 60 min.

4.2 *Drop Test*—Perform the drop test in accordance with Method [D775](#).

4.2.1 Determine the drop height of the specimen from the following:

Gross Weight, lb (kg)	Drop Height, in. (mm)
0 to 20 lb (0 to 9.1 kg)	16 in. (406 mm)
20 to 30 lb (9.1 to 13.6 kg)	14 in. (356 mm)
30 lb (13.6 kg) and over	12 in. (305 mm)

4.2.2 Refer to Fig. A1.1 of Test Method [D5276](#) and perform the following drop sequence: one flatwise drop onto the top (Face 1); one flatwise drop onto the bottom (Face 3); one drop onto each of the top edges (edges 1-2, 1-5, 1-4, and 1-6), and one drop onto each of the bottom edges (edges 3-2, 3-5, 3-4, and 3-6). Complete one drop sequence (10 drops).

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Apparatus.

Current edition approved Nov. 1, 2013. Published November 2013. Originally approved in 1983. Last previous edition approved in 2008 as E920 – 97 (2008). DOI: 10.1520/E0920-97R13.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

5. Acceptance Criteria

5.1 *Disposable Glass*—Completion of testing without breakage occurring in any of the three units tested. In the case of bulk packed items (more than 50 in a shipping container), less than 3 % total breakage in each of the three units tested.

5.2 *All Other*—Completion of testing without breakage occurring in any of the three units tested.

5.3 Outer package must contain product and provide a degree of protection to contents. One complete retest consisting of three new packages is permitted.

6. Report

6.1 Report results of test as specified in Method D775.

6.2 Each supplier shall maintain on file laboratory test reports verifying successful testing as outlined in this specification, and if required, issue a written certification stating that packaging complies with this specification.

7. Keywords

7.1 apparatus; commercial; laboratory; packaging

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT).