INTERNATIONAL ELECTROTECHNICAL COMMISSION

IEC TR 61000-1-6 Edition 1.0 2012-07

Electromagnetic compatibility (EMC) -

Part 1-6: General – Guide to the assessment of measurement uncertainty

CORRIGENDUM 1

Figure 17 - Situations, where and how an instrument is suitable for tests and/or measurements as specified in the applicable basic standard with tolerances

Replace the text of the existing footnote 1) with the following new text:

1) This approach deviates from other concepts found in literature, because the tolerances given in the standards of the 61000-4-x series once were defined without taking into account uncertainties. Therefore uncertainties should not be used to demonstrate compliance with tolerances.