TECHNICAL SPECIFICATION

IEC TS 61895

First edition 1999-10

Ultrasonics – Pulsed Doppler diagnostic systems – Test procedures to determine performance

Ultrasons – Systèmes de diagnostic à effet Doppler pulsés – Procédures d'essai pour déterminer la performance



Reference number IEC/TS 61895:1999(E)

Numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series.

Consolidated publications

Consolidated versions of some IEC publications including amendments are available. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

Validity of this publication

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology.

Information relating to the date of the reconfirmation of the publication is available in the IEC catalogue.

Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is to be found at the following IEC sources:

- IEC web site*
- Catalogue of IEC publications Published yearly with regular updates (On-line catalogue)*
- IEC Bulletin Available both at the IEC web site* and as a printed periodical

Terminology, graphical and letter symbols

For general terminology, readers are referred to IEC 60050: International Electrotechnical Vocabulary (IEV).

For graphical symbols, and letter symbols and signs approved by the IEC for general use, readers are referred to publications IEC 60027: *Letter symbols to be used in electrical technology*, IEC 60417: *Graphical symbols for use on equipment. Index, survey and compilation of the single sheets* and IEC 60617: *Graphical symbols for diagrams.*

* See web site address on title page.

TECHNICAL SPECIFICATION

IEC TS 61895

First edition 1999-10

Ultrasonics – Pulsed Doppler diagnostic systems – Test procedures to determine performance

Ultrasons – Systèmes de diagnostic à effet Doppler pulsés – Procédures d'essai pour déterminer la performance

© IEC 1999 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission3, rue de Varembé Geneva, SwitzerlandTelefax: +41 22 919 0300e-mail: inmail@iec.chIEC web site http://www.iec.ch



Commission Electrotechnique Internationale International Electrotechnical Commission Международная Электротехническая Комиссия



For price, see current catalogue

V

CONTENTS

Page

FOREWORD	4
INTRODUCTION	6

Clause

1	Scop	е		7
2	Norm	ative re	ferences	7
3	Defin	itions		7
4	Svmb	ols		12
5	Over	all tests	of complete systems	12
0	5 1	Genera	al considerations	12
	0.1	5 1 1	Types of pulsed Doppler ultrasound systems	12
		512	Worst-case conditions	13
		5.1.3	Doppler beam axes	14
		5.1.4	Probe/target distance variation and measurement	14
	5.2	Initial c	conditions	15
		5.2.1	Power supply	15
		5.2.2	Target movement direction	15
		5.2.3	Propagation medium	15
		5.2.4	Penetration depth	16
		5.2.5	Working depth	16
		5.2.6	Focusing	16
		5.2.7	Working Doppler angle	16
		5.2.8	Wall-thump filter cut-off frequency	17
		5.2.9	Transmitter output power	17
		5.2.10	Working pulse repetition frequency (PRF)	17
		5.2.11	Doppler (receiver) gain	17
		5.2.12	Test frequency	17
		5.2.13	Working sample volume length	17
		5.2.14	Doppler signal power measurement	17
	5.3	Zero si	gnal noise level	17
	5.4	Dopple	er frequency response	18
		5.4.1	Frequency response range	18
		5.4.2	Deviation from flat response	18
		5.4.3	Large signal performance	18
	5.5	Spatial	response	19
		5.5.1	Sample volume response	19
	5.6	Sampl	e volume position registration error	20
	5.7	Beam _I	position and orientation	21
	5.8	Intrinsi	c broadening	22
	5.9	Dead z	cone	22
	5.10	Acoust	ic working frequency	22
	5.11	Flow di	irection separation	22
	5.12	Velocit	y estimation accuracy	22

	5.13	Volume flow estimation accuracy	22
	5.14	Maximum, mean, mode and median frequency estimation accuracy	23
	5.15	Velocity waveform indices estimation accuracy	25
6	Dopp	ler test objects	26
	6.1	Test objects	26
	6.2	Electronic test object	26
Ann	ex A ((normative) Description of pulsed Doppler ultrasound systems	28
A.1	Sing	gle-channel system	28
A.2	Mult	ii-channel system	29
A.3	Alia	sing	29
A.4	Dup	lex and triplex scanners	29
Bibl	iograp	bhy	32

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ULTRASONICS – PULSED DOPPLER DIAGNOSTIC SYSTEMS – TEST PROCEDURES TO DETERMINE PERFORMANCE

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this technical specification may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. In exceptional circumstances, a technical committee may propose the publication of a technical specification when

- the required support cannot be obtained for the publication of an International Standard, despite repeated efforts, or
- the subject is still under technical development or where, for any other reason, there is the future but no immediate possibility of an agreement on an International Standard.

Technical specifications are subject to review within three years of publication to decide whether they can be transformed into International Standards.

IEC 61895, which is a technical specification, has been prepared by IEC technical committee 87: Ultrasonics.

The text of this technical specification is based on the following documents:

Enquiry draft	Report on voting
87/151/CDV	87/168/RVC

Full information on the voting for the approval of this technical specification can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 3.

Annex A forms an integral part of this technical specification.

The committee has decided that this publication remains valid until 2005. At this date, in accordance with the committee's decision, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

INTRODUCTION

Pulsed ultrasonic Doppler flowmeters and velocimeters are widely used in clinical practice, usually in combination with real-time **B-mode** imaging and colour-flow imaging instruments. The device periodically transmits pulses of ultrasound from an ultrasound transducer and measures the Doppler shift in the frequency of ultrasound reflected and scattered from moving tissues. This Doppler shift is proportional to the component of reflector or scatterer velocity along the ultrasound beam. By looking for Doppler shifts in the received signal at specific times after transmission (range-gating), the device can be used to determine the variation of tissue velocity with distance along the ultrasound beam. The device is sensitive to movement only within a region of the beam called the sample volume. The position of the sample volume along the beam may be adjusted by altering the delay between transmission and range-gating. Multi-channel devices have a number of sample volumes operating simultaneously.

The pulsed ultrasonic device is most commonly used to investigate blood flow when the ultrasound is scattered from red blood cells.

This technical specification describes a range of tests which may be used to measure performance and the test objects required. In many cases, the test method and test object have been described in IEC 61206 and in these cases reference is simply made to this document. Other tests and test objects are described in [1] and [2]. The test methods may be considered as falling into one of the following three categories. The first is routine quality control tests that can be carried out by a clinician or technologist to ensure that the system is working adequately or has adequate sensitivity. The second is more elaborate test methods, conducted less frequently, when, for example, the system is suspected of malfunctioning. The third represents tests that would be carried out by a manufacturer on complete systems in order to guarantee compliance with specification.

ULTRASONICS – PULSED DOPPLER DIAGNOSTIC SYSTEMS – TEST PROCEDURES TO DETERMINE PERFORMANCE

1 Scope

This technical specification describes

- test methods for measuring the performance of pulsed **Doppler ultrasound systems**;
- **Doppler test objects** for carrying out these tests;

and applies to

- tests made on an overall pulsed **Doppler ultrasound system**, a system which is not disassembled or disconnected;
- tests made on pulsed **Doppler ultrasound systems** whether they are stand-alone or as part of another ultrasound instrument.

Electrical safety, acoustic output and electromagnetic compatibility (EMC) are not covered in this technical specification.

The workload to perform all described tests is, in general, prohibitive. It is intended that a subset of the described tests is adopted for regular use. However, experience to give guidance for selection has still to be gathered and will be the subject of ongoing work.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this technical specification. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this technical specification are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of IEC and ISO maintain registers of currently valid International Standards.

IEC 61102:1991, Measurement and characterisation of ultrasonic fields using hydrophones in the frequency range 0,5 MHz to 15 MHz

IEC 61206:1993, Ultrasonics – Continuous-wave Doppler systems – Test procedures

IEC 61390:1996, Ultrasonics – Real-time pulse-echo systems – Test procedures to determine performance specifications

3 Definitions

For the purposes of this technical report, the following definitions apply.

3.1

6 dB spectral width

width of a frequency spectrum between the frequencies at which the spectral power is 6 dB less than the maximum power

20 dB spectral width

width of a frequency spectrum between the frequencies at which the spectral power is 20 dB less than the maximum power

3.3

acoustic working frequency

centre frequency

zero-crossing acoustic-working frequency of the transmitted pulse spectrum

[3.4 of IEC 61102, modified]

3.4

aliasing

false indication of signal frequency as a result of sampling at too low a frequency

NOTE Aliasing occurs when the Doppler frequency exceeds the Nyquist limit frequency of the Doppler ultrasound system. In a non-directional system, the indicated frequency of the Doppler signal is the true Doppler frequency mirrored in the Nyquist limit frequency. In a directional system, the indicated frequency of the Doppler signal is the true Doppler frequency mirrored in the Nyquist limit frequency mirrored in the Nyquist limit frequency mirrored in the Nyquist limit frequency and changed in sign. In systems using a baseline shift, the term Doppler frequency should be replaced by Doppler frequency plus baseline frequency shift in the above explanation.

3.5

baseline frequency shift

frequency by which the Doppler signal is shifted before analysis in order to alleviate the effects of aliasing

3.6

B-mode (brightness-modulated display)

image generated by a pulse-echo ultrasound scanner in which the echoes from reflectors and scatterers in the tissues swept by a pulsed ultrasound beam are represented by a brightness-modulated two-dimensional display

3.7

clutter

unwanted components of the Doppler signal as it appears after the **Doppler demodulator**

NOTE **Clutter** arises from stationary or slowly moving reflectors and is usually removed by high-pass filters (wall-thump filters) within the **Doppler ultrasound system**.

3.8

dead zone

region close to the transducer in which the **system** is insensitive to tissue movement

3.9

directional

direction sensing

descriptor of a type of **Doppler ultrasound system** which indicates whether scatterers or reflectors are approaching or receding from the ultrasonic transducer

[1.3.1 of IEC 61206]

3.10

direction resolving

direction separating

descriptor of a type of **Doppler ultrasound system** in which the **Doppler output** appears at different output terminals, **output channels** or **output devices** depending on the direction of scatterer or reflector motion relative to the transducer

[1.3.2 of IEC 61206, modified]

Doppler angle

acute angle between the axis of the ultrasound beam during Doppler measurements and the direction of movement of the scatterer or reflector

3.12

Doppler demodulator

that part of the **Doppler ultrasound system** at which the Doppler signal is derived through mixing of the received signal and a **reference signal**

3.13

Doppler frequency

Doppler-shift frequency

change in frequency of an ultrasound scattered or reflected wave caused by relative motion between the scatterer or reflector and the transducer. It is the difference in frequency between the transmitted and the received wave

[1.3.3 of IEC 61206, modified]

3.14

Doppler output connector

electrical connector or that part of the **Doppler ultrasound system** at which the **Doppler output** is available for connection to external **output devices**

[1.3.5 of IEC 61206]

3.15

Doppler output

signal at the Doppler frequency or at Doppler frequencies which activates the output device [1.3.4 of IEC 61206, modified]

3.16

Doppler spectrum

set of Doppler frequencies produced by a Doppler ultrasound system

[1.3.6 of IEC 61206]

3.17

Doppler test object

artificial structure used in testing Doppler ultrasound systems

3.18

Doppler ultrasound system

equipment designed to transmit and receive ultrasound and to generate a **Doppler output** from the difference in frequency between the transmitted and received waves

[1.3.8 of IEC 61206]

3.19

duplex scanner

ultrasound instrument which combines real-time **B-mode** imaging with a **Doppler ultrasound** system

3.20

nominal Doppler beam direction axis

assumed axis of the ultrasonic beam from the transducer used for Doppler measurements. This axis will often be the axis of rotational symmetry of the Doppler probe for a single-element transducer

nominal first lateral Doppler beam axis

coordinate axis perpendicular to the **nominal Doppler beam direction axis** and with a position indicated on the probe body for a single-beam direction probe or contained within the scan plane for the probe of a **duplex** or **triplex scanner** (see figure A.2)

3.22

nominal second lateral Doppler beam axis

coordinate axis perpendicular to both the **nominal Doppler beam direction axis** and the **nominal first lateral Doppler beam axis.** This axis is perpendicular to the scan plane for a the probe of a **duplex** or **triplex scanner** (see figure A.2)

NOTE The nominal first lateral Doppler beam axis, the nominal second Doppler beam axis and the nominal Doppler beam direction axis form a right-handed Cartesian co-ordinate set as shown in figure A.2.

3.23

nominal sample volume length

length of the sample volume indicated by the system. This normally will be a numerical display or a distance between markers on a screen indicating the extent of the sample volume along the nominal Doppler beam direction axis

3.24

non-directional

descriptor of a type of Doppler ultrasound system which is not directional

[1.3.9 of IEC 61206]

3.25

Nyquist limit frequency

half the **pulse repetition frequency.** In systems not using a baseline shift, it equals the frequency under which **aliasing** does not occur

3.26

observed velocity

component of the velocity of a scatterer or reflector along the axis of the ultrasound beam. This is directed towards or away from the transducers

3.27

output channel

part of the **Doppler ultrasound system** which functionally represents a particular aspect of the **Doppler output**

NOTE A Doppler ultrasound system may have two output channels, each representing a flow in a particular direction.

[1.3.12 of IEC 61206]

3.28

output device

any device included in a **Doppler ultrasound system** or capable of being connected to it that makes the **Doppler output** accessible to the human senses

[1.3.13 of IEC 61206]

3.29

phase-quadrature demodulation

a method of derivation of Doppler signals incorporating flow direction information, in which two **Doppler demodulators** are used with **reference signals** 90° out of phase – leading to inphase and quadrature Doppler signals 90° out of phase. The direction of the phase shift between the in-phase and quadrature parts of the Doppler signal component at a particular frequency indicates the direction of movement of the **target** giving rise to that component

pulse repetition frequency (PRF)

the number of pulses or bursts of ultrasound emitted by the transducer per second

3.31

range gate

that part of the **Doppler ultrasound system** which selects signals received from a range of depths to generate the Doppler signal. This is achieved by selecting signals arriving during a time interval following a delay after pulse transmission

3.32

reference signal

the signal mixed with the received signal in the **Doppler demodulator** in order to generate the Doppler signal. In **non-directional Doppler ultrasound systems** or **systems** using **phase-quadrature demodulation** the **reference signal** is at the transmitted frequency. In off-set frequency **directional Doppler ultrasound systems**, the **reference signal** is at the transmitted frequency plus or minus the off-set frequency

3.33

sample volume

the region of the ultrasound beam in which moving scatterers or reflectors give rise to components of the Doppler signal from the **Doppler ultrasound system**

3.34

scan plane

plane containing ultrasonic scan lines

[3.29 of IEC 61390]

3.35

sonogram

a frequency-time display in which the relative amplitude or power of each frequency component of the detected signal from successive or overlapping time windows is displayed as adjoining vertical grey scale lines

3.36

spectral width

range of Doppler frequencies within the Doppler spectrum

3.37

spectrum

a display of amplitude or power against frequency, showing the relative amplitude or power of each frequency component contained in the detected signal (see **Doppler spectrum**)

3.38

string test object

line of scatterers moving at a constant velocity in the direction of the line, and having ultrasound scattering properties similar to that of a moving column of blood

3.39 system Doppler ultrasound system See 3.18

target

reflector, scatterer or collection of scatterers giving rise to a received signal

3.41

target depth

distance from the probe face to the target along the beam

3.42

triplex scanner

ultrasound instrument which combines real-time **B-mode** imaging and colour-flow imaging with a pulsed **Doppler ultrasound system**

- 12 -

4 Symbols

- c = average speed of sound in a medium
- *L* = distance from the transducer face to the centre of the **sample volume**
- *L*_{max} = maximum depth of penetration of the pulsed **Doppler ultrasound system**

*f*₀ = acoustic working frequency

*f*_{vm} = frequency of sinusoidal variation in mean frequency of simulated Doppler signal used in maximum, mean, mode and median frequency estimation accuracy test

5 Overall tests of complete systems

5.1 General considerations

5.1.1 Types of pulsed Doppler ultrasound systems

A pulsed **Doppler ultrasound system** may be **directional**, **non-directional** or **directionresolving**. **Directional** (or directional sensing) refers to a type of pulsed **Doppler ultrasound system** which indicates whether reflectors or scatterers are approaching or receding from the ultrasonic transducer. **Non-directional Doppler ultrasound systems** do not indicate the direction of movement. **Direction-resolving** (or direction-separating) **Doppler ultrasound systems** provide for **Doppler outputs** to appear at different **output channels** depending on the direction of reflector or scatterer movement. The **system** may use a **phase-quadrature demodulation** or offset reference frequency demodulation in order to derive Doppler signals retaining flow direction information. Annex A gives descriptions and diagrams of these different types of pulsed **Doppler ultrasound systems**.

The **system** may be a stand-alone instrument or part of a **B-mode** and/or flow imaging **system**. The stand-alone instrument may make use of a single transducer for transmission and reception or separate transducers for these functions, in which case the instrument may be switched to operate in the continuous wave mode. When incorporated with a **B-mode** real-time imaging instrument, a separate transducer may be used for pulsed Doppler operation or the same transducer used for pulse echo imaging and pulsed Doppler work.

The system may be part of a duplex or a triplex scanner. Duplex scanners allow display of the nominal Doppler beam axis direction used for Doppler measurements to be displayed on the **B-mode** image along with indications of the depth and length of the sample volume. Provision is made for the operator to line up an electronic marker parallel to the axis of the displayed blood vessel in order that the instrument may calculate the angle between the ultrasound beam and the direction of the vessel. This allows the conversion of **Doppler frequencies** to blood velocities on the assumption of flow in the axial direction. Triplex scanners, in addition to the functions of the duplex scanners, display images of moving blood, colour-coded according to blood velocity, superimposed on the B-mode image.

The **system** may or may not be equipped with automatic adaptation of operating parameters to the depth of the **sample volume** and the nature of the tissue between the transducer and **sample volume**. Examples of parameters which are adapted in this way are **pulse repetition** frequency (PRF), focal depth, transducer aperture and transmission signal spectrum.

The **system** may incorporate a method of spectral analysis of the Doppler signal, displaying the time-varying frequency spectrum of the Doppler signal. This frequency analysis may be based on the Fast Fourier Transform (FFT) or other methods of spectral analysis. The **system** may as an alternative display the time-varying maximum, mean, mode or median **Doppler** frequency derived from the spectral analyser, or more directly, by time-domain processing.

The **system** may incorporate interactive or automated measurement and/or calculation systems to process further the data from the spectrum analysis and/or **Doppler frequency** waveform – calculating, for example, indices of waveform shape and **spectral width**.

The **system** may incorporate means for the operator to listen to the Doppler signal using a loudspeaker or headphones.

The **system** may be a multi-channel instrument having a number of **sample volumes** and associated Doppler signal channels.

5.1.2 Worst-case conditions

A test method may be applied to determine a particular performance parameter of a **system**. Often a number of quantities can have a bearing on overall performance, each of which requires the application of a distinct test method. Some of these quantities need to be maximized and others need to be minimized in order to obtain best overall performance. Considering overall performance, table 1 gives the worst-case conditions for key quantities appropriate to pulsed **Doppler ultrasound systems** and the corresponding subclause numbers which describe a suitable test method. As an example, if the penetration as mentioned in 5.2.4 is minimized, this will lead to worst-case overall performance; conversely, maximizing penetration will lead to maximized performance.

Worst case is minimum value of		Worst case is maximum value of		
Quantities Subclauses		Quantities	Subclauses	
Penetration depth	5.2.4	Dead zone	5.9	
Clutter rejection index	5.4.3.1	High-frequency response error	5.4.1	
Flow direction separation	5.11	Low-frequency response error	5.4.1	
		Harmonic distortion	5.4.3.2	
		Intermodulation distortion	5.4.3.3	
		Sample volume registration error	5.6	
		Intrinsic broadening	5.8	
		Beam position and orientation error	5.7	
		Velocity estimation error	5.12	
		Volume flow estimation error	5.13	
		Maximum, mean, mode and median frequency estimation errors	5.14	

Table 1 – Worst case for various quantities, and corresponding subclause numbers

5.1.3 Doppler beam axes

For the purposes of determining the orientation of the ultrasound probe, three orthogonal axes should be defined. One of these is the **nominal Doppler beam direction axis** and the other two are the first and second lateral Doppler beam axes. The first lateral axis is perpendicular to the sound beam axis and is in the plane of the scan. The second lateral axis is also perpendicular to the sound beam axis but is perpendicular to the plane of the scan. In a self-contained pulsed **Doppler ultrasound system**, these axes should be defined with respect to a mark or feature on the probe body. In a **duplex** or **triplex scanner**, the first lateral Doppler beam axis should be within the **scan plane** (see annex A).

5.1.4 Probe/target distance variation and measurement

The probe of the **Doppler ultrasound system** should be attached to a calibrated positioning mechanism capable of movement in three orthogonal directions. The directions of movement should be parallel to the **nominal Doppler beam direction axis** and parallel to the **nominal first** and **second lateral Doppler beam axes**. Alternatively, the probe of the **Doppler ultrasound system** may be held in a fixed position and the **target** moved parallel to these axes.

5.2 Initial conditions

These subclauses describe conditions common to all of the tests given in 5.2.4 and 5.3 to 5.15, as well as a procedure to determine the appropriate **Doppler frequency** and distance ranges to be used for these measurements.

Where a particular type of **system** may be comprised of various combinations of components, it is intended that each combination should be regarded as a separate **system** for testing purposes. For example, a **system** may have various transducer options. In this case, each transducer and output recording or presentation device connected to the basic **system** will define a different **system**. For the test to be meaningful, all instrument control settings should be recorded during the test. For those tests where the operator is required to make an assessment of performance from a grey scale or colour video display, the video monitor settings and ambient lighting conditions should be such that the lowest displayed intensity level above background is clearly distinguishable from the background.

5.2.1 Power supply

To ensure that the stated specifications hold over the range of power supply voltage, tests should be undertaken for the different power line voltages and worst-case test result values reported. The power line voltages are to be used at their nominal values and at 10 % above and below the nominal voltage. For power line operated **systems**, the worst-case values are those obtained after a specified warm-up time.

Portable, battery-operated **systems** weighing less than one kilogram should be tested with no warm-up and only over the time span sufficient to perform each test to simulate typical use. Heavier battery-powered **systems** should be tested under the same conditions as the power line operated **systems**.

For all battery-operated **systems** the results should be the worst case found over the span of battery voltages from the fully charged condition to a nominal end-of-life voltage. Any **system** tuning or adjustment should be carried out as specified in the instructions supplied to the user. It should be stated whether the nominal life-span of the battery occurs under continuous or intermittent conditions of use. This allows the manufacturer to select the intended normal battery life for either intermittent or continuous use. Manufacturers may report best-case results (for example, with fully charged batteries) in addition to worst-case results, provided that they report the **system** condition under which the results were obtained.

5.2.2 Target movement direction

If the pulsed **Doppler ultrasound system** is part of a **duplex** or **triplex scanner** then the **target** movement should be within the **scan plane** of the imaging system, unless otherwise stated. If the pulsed **Doppler ultrasound system** is stand-alone, then the **target**'s movement should be within the plane defined by the **nominal Doppler beam direction axis** and the **nominal first lateral Doppler beam axis**, unless otherwise stated.

5.2.3 **Propagation medium**

Most of the tests listed below can be performed in an appropriate non-attenuative medium whose acoustic velocity is 1 540 ms⁻¹, such as a 9,0 % glycerol solution by volume. If an electronic injection system is used, this may be coupled to the transducer under test using a solid medium such as perspex which has a higher acoustic velocity of 2 700 to 2 800 ms⁻¹. Attenuation using a suitable absorbor is required for the measurement of penetration depth. This should have an acoustic velocity of 1 540 ms⁻¹. The attenuation may be tissue equivalent (0,45 dB cm⁻¹ MHz⁻¹ to 0,55 dB cm⁻¹ MHz⁻¹), as commonly used in a flow phantom, or it may be higher, as it may be necessary to perform sensitivity measurements using an attenuative polyurethane wedge in conjunction with a string test object.

5.2.4 Penetration depth

The maximum depth of penetration along the ultrasound beam (L_{max}) is measured using a **Doppler test object** with an attenuating tissue-mimicking medium between the probe and **target**. The **target depth** is increased until the Doppler signal power is equal to the noise power – that is, when the power of the **Doppler output** with a moving **target** (signal plus noise) is double that with a stationary **target** (noise only). In this case the signal-to-noise ratio (SNR) is zero decibels.

If the maximum depth in the **Doppler test object** is insufficient to reduce the signal power to give a SNR of zero decibels, the transmitter power should be reduced to enable the above condition to be achieved. In this case, the transmitter output setting should be noted.

The attenuation coefficient of the tissue mimicking medium should be within the range 0,45 dB cm^{-1} MHz⁻¹ to 0,55 dB cm^{-1} MHz⁻¹, as commonly used in a flow phantom, or it may be higher (for example 0,70 dB cm^{-1} MHz⁻¹ to 0,80 dB cm^{-1} MHz⁻¹) and should be reported along with the results of this test.

It should be noted that penetration depends on the target, and comparisons between **systems** are valid only if similar targets are used.

5.2.5 Working depth

Except where otherwise stated, measurements should be made at a distance between the probe face and the **target** (along the beam) of $L_{max}/2$, where L_{max} is the maximum depth of penetration (see 5.2.4).

5.2.6 Focusing

When testing **systems** with an operator-set (non-automatic) variable focus in **duplex** or **triplex scanners**, the nominal focus should be adjusted to the depth of the centre of the **sample volume**, or as close as the **system** focus and **sample volume** depth settings allow.

5.2.7 Working Doppler angle

Except where stated otherwise, the angle between the **nominal Doppler beam direction axis** and the direction of **target** movement in the **Doppler test object** should be 0°, 30°, 45°, or 60°. It should be noted that it may not be possible to achieve all these angles in all test objects (for example 0° in most flow test objects).

In a stand-alone pulsed **Doppler ultrasound system** with a transducer mounted in such a way that the beam axis is designed to be coincident with the probe axis, the probe axis should be taken as the **nominal Doppler beam direction axis**. In other isolated **systems**, the **nominal Doppler beam direction axis** should be described by the manufacturer.

In **duplex** and **triplex scanners**, the working Doppler angle is that measured between the Doppler beam and the direction of **target** movement as indicated on the image screen. If the Doppler beam orientation is variable, this angle should be set to 60°.

5.2.8 Wall-thump filter cut-off frequency

Except where otherwise stated, in **systems** with variable wall-thump filter cut-off frequencies this frequency should be set to $4 \times 10^{-5} f_0$ or the nearest value allowed by the **system** for peripheral vascular applications, where f_0 is the **acoustic-working frequency**. This corresponds approximately to 3 cms⁻¹ for a **Doppler angle** of zero. For other systems other filter cut-off frequencies may be more appropriate. For example, for fetal applications this frequency should be set to $2 \times 10^{-5} f_0$ or the nearest value allowed by the **system**, and for adult cardiology applications, this frequency should be set to $6 \times 10^{-5} f_0$ or the nearest value allowed by the **system**.

5.2.9 Transmitter output power

Except where otherwise stated, in **systems** where the transmitted output power is variable, this power should be set at maximum.

5.2.10 Working pulse repetition frequency (PRF)

Except where otherwise stated, the PRF should be set to $0.4c/L_{max}$ or the nearest lower frequency allowed by the **system**.

5.2.11 Doppler (receiver) gain

Except where otherwise stated, in **systems** with a real-time spectrum analyser as an **output device**, the Doppler gain should be set in such a way that, with the transmitter power at minimum and the **target** stationary, the noise is just not visible on the display. Alternatively, in **systems** with no spectral analyser, the Doppler gain should be set to give a small but measurable noise signal on the **Doppler output**.

5.2.12 Test frequency

The test frequency is the **Doppler frequency** to be used in performing the tests. For the electronic injection test device, this is the frequency of the injected audio signal. For other test devices, this is the mean **Doppler frequency**. Except where otherwise stated, a test frequency of $5 \times 10^{-4} f_0$ or a frequency specified by the manufacturer should be used.

5.2.13 Working sample volume length

Except where otherwise stated, in **systems** where the **sample volume** length is variable, the **nominal sample volume length** should be set to $25 \times 10^6/f_0$ mm, or the nearest greater length allowed by the **system**, or if a greater length is not available, the nearest length, where f_0 is the **acoustic working frequency** in hertz.

5.2.14 Doppler signal power measurement

Doppler signal power can be measured at a **Doppler output connector**. The Doppler signal power may be measured by any device having an accuracy of better than 5 % at the signal level under consideration. It should be quoted as mean square signal volts, root-mean-square (r.m.s.) volts or in decibels relative to 1,0 volts r.m.s. A true reading r.m.s. meter is suitable. Note that only relative values are required in all the tests.

5.3 Zero signal noise level

With the test object producing a Doppler signal at the test frequency, the receiver level and transmitted power output are set to display the Doppler signal on the **sonogram** such that the full grey scale range of the display is utilized. The motion of the target is stopped, or in the case of the electronic injection device, the injected signal is switched off. The signal level at the **Doppler output** is noted as the zero signal noise level.

5.4 Doppler frequency response

The **Doppler frequency** response and accuracy are preferably tested with a test object giving rise to a narrow **Doppler spectrum**. An electronic **Doppler test object** with a single-frequency Doppler shift or a string **Doppler test object** may be used.

5.4.1 Frequency response range

The baseline of the frequency display should be set to zero. Frequency response is measured for the frequency range from 0 Hz to the upper Nyquist limit. The shift frequency of the electronic **Doppler test object** or the speed of the moving **target** in a string or flow **Doppler test object** is varied over this range. The time-average **Doppler output** signal level is measured as a function of **Doppler frequency** using an r.m.s. voltmeter or power meter and a means of mean frequency measurement. If the **Doppler output** signal level has one peak value, the low-frequency response frequency and the high-frequency response frequency are found from those frequencies at which the output voltage is 3 dB less than its peak level, although other limits may be used if so declared. The same procedure should apply in the case of multiply peaked response curves, where the lowest value between the peaks is above the -3 dB level relative to the largest peak described above.

If the response curve is multiply peaked and the lowest value between the peaks is below the -3 dB level relative to the largest peak described above, then the smallest value found between the peaks should be taken as the minimal detectable signal level. A horizontal line graph at this signal will then intercept the frequency response curve at this minimum and two other points. These two other points are the low- and high-frequency response values and should be quoted as the result of the test, qualified by a statement of the level of this minimum relative to the highest value.

In all of the above cases the minimum and maximum frequency limits should be compared with those indicated by the **system** or quoted in the **system** specification and the errors reported. Note that the minimum frequency is the wall-thump cut-off frequency.

The procedure should be repeated for the negative frequency range, from 0 Hz to the lower Nyquist limit.

5.4.2 Deviation from flat response

For the frequency range from zero to the upper Nyquist frequency, half the range of the signal level (in decibels), from the level at the maximum and minimum frequency limits to the maximum level measured, should be quoted as the maximum deviation from a flat response.

The procedure should be repeated for the negative frequency range, from 0 Hz to the lower Nyquist limit.

5.4.3 Large signal performance

Large spurious signals can cause errors in ultrasound Doppler receivers which behave similarly to communication system receivers. The tests in this section look for the magnitude of these effects for interfering signals that are of the order of the maximum levels that will be encountered in practice.

5.4.3.1 Fixed target effect on sensitivity (clutter rejection)

The effect of highly reflecting fixed **targets** on the amplitude of the **Doppler output** can be determined by using the small vessel or string **Doppler test object**. The change in the **Doppler output** should be reported in terms of the change (in decibels) observed when a highly reflecting **target** is placed within the **sample volume** in the far side of the string or small vessel from the transducer. The reflecting **target** should be placed as close as practicable to the moving string or vessel and oriented to produce the maximum fixed **target** echo. A spherical **target** large enough to produce a strong echo without overload is suggested in order to reduce alignment problems. The **Doppler output** should be observed while moving the fixed **target** laterally and axially within the **sample volume** and the maximum change in **Doppler output** reported.

Alternatively, the electronic **Doppler test object** may be used to determine the instrument's **clutter** rejection. The electronic **Doppler test object** is operated such that it generates a test frequency at such an amplitude that the **Doppler output** is at least 20 dB greater than the noise level but significantly below saturation. The **Doppler test object** is then operated in such a way that it generates simultaneously a stationary (zero Doppler shift) frequency signal and the amplitude of the signal is increased until the **Doppler output** changes by 3 dB. Here, it is assumed that the signal at the zero **Doppler frequency** does not contribute directly to the Doppler ouput power. The ratio of the stationary to test **Doppler frequency** signal amplitudes generated by the electronic **Doppler test object** is recorded as the **clutter** rejection index.

5.4.3.2 Harmonic distortion

This may be measured as detailed in 2.3.3.1 of IEC 61206. Alternatively, an electronic **Doppler test object** may be used to measure harmonic distortion. Using a spectrum analyser connected to the Doppler output connector, the sum of the powers of the harmonics of the test frequency is measured and expressed as a percentage of the signal power at the test frequency. This is the percentage harmonic distortion. The distortion is measured and reported at Doppler signal output levels at the test frequency of 10 dB above the noise level and at additional 10 dB increments until the distortion exceeds 10 %.

5.4.3.3 Intermodulation distortion

This may be measured as detailed in 2.3.3.3 of IEC 61206. Alternatively, an electronic **Doppler test object** may be used to measure harmonic distortion. The electronic **Doppler test object** is operated such that it generates a low-frequency signal at a frequency of 10 % of the test frequency in addition to a signal at the test frequency. The low-frequency signal level should be maintained at 30 dB above the test frequency signal. Using a spectrum analyser connected to the Doppler output connector, the sum of the powers of the signals at the test frequency plus and minus the low frequency is measured and expressed as a percentage of the signal power at the test frequency. This is the percentage intermodulation distortion. The distortion is measured and reported at Doppler signal output levels at the test frequency of 10 dB above the noise level and at additional 10 dB increments until the distortion exceeds 10 %.

5.5 Spatial response

The relative sensitivity of the **Doppler ultrasound system** to scatterers at different points in space can be determined by these procedures. The measurements are made at the nominal first lateral Doppler beam axis orientation unless otherwise stated.

5.5.1 Sample volume response

The following tests can be used to measure the sensitivity variation within the **sample volume** and its lateral widths and axial length. These distances are those between points at which the instrument response to a moving **target** falls to a specified fraction of the maximum response.

The test should be carried out at the working **sample volume** length setting at depths of 0,2 kL_{max} where k = 1, 2, 3, 4, or as specified by the manufacturer.

It should be noted that measurements made with different **targets** are likely to give rise to different values for the dimensions of the **sample volume** since the sensitivity variation depends on **target** acoustic response. The measurements using one type of **target** may, nevertheless, be used for comparative purposes.

In particular, the Doppler signal power when using an **ideal string test object** would be proportional to the line integral of the sensitivity variation (signal power from a moving single monopole scatterer versus position) along the string path. Non-ideal behaviour in real-string **Doppler test objects** resulting from finite scatterer dimensions, string thickness and periodicity in string structure, for example, has yet to be fully investigated [2].

As a consequence of the complex nature of the scatter from small-sphere **targets** [3], and in particular its variation with frequency, the small-ball **Doppler test object** is not recommended for use in pulsed **Doppler ultrasound systems** or **systems** with narrow beam widths until its performance limits have been evaluated.

NOTE When measuring low Doppler signal power when the background noise is not negligible (for example, near L_{MAX}), then the measured noise power should be subtracted from the signal-plus-noise power measured at the Doppler signal connector in order to estimate the Doppler signal power.

5.5.1.1 Sample volume axial response

With the Doppler beam direction axis set at the working **Doppler angle** with respect to the string, the probe is moved axially and laterally while monitoring the Doppler signal power until this is a maximum. The probe is then moved parallel to the **nominal Doppler beam direction axis** and the distance between the points at which the Doppler signal power falls to -6 dB of the maximum and the distance between the points at which the power falls to -20 dB of the maximum are noted and reported as the 6 dB and 20 dB axial **sample volume** lengths at 60°. These measurements are repeated at **Doppler angles** of 45° and 75° and the measurements extrapolated to find the **sample volume** length at 90° . This length should be compared with the **nominal sample volume length** indicated by the pulsed **Doppler ultrasound system**.

With the centre of the **sample volume** at the working depth, the length should be altered from its minimum to maximum value in five equal steps or those allowed by the **system**, and the 6 dB and 20 dB **sample volume** lengths at 90° reported together with the **nominal sample volume lengths** indicated by the pulsed **Doppler ultrasound system**.

5.5.1.2 Sample volume lateral responses

The measurements carried out in 5.5.1.1 are repeated, with the difference that the probe is moved laterally perpendicular to the string. This gives the first lateral sample volume width. The measurements are repeated with the probe rotated by 90° about the **nominal Doppler beam direction axis**, giving the second lateral sample volume width.

5.6 Sample volume position registration error

Using a string **Doppler test object**, the **sample volume** is moved along the beam until the Doppler signal power is at maximum and this power is noted. The **sample volume** is then moved to the two positions at which the Doppler signal power is 3 dB lower than the maximum. With a **duplex** or **triplex scanner** the distance along the Doppler beam between the point halfway between the centres of the **sample volumes** (as indicated on the screen) in these positions and the image of the string is reported as the registration error.

5.7 Beam position and orientation

These tests are designed to measure the deviation of beam axis position and orientation with respect to the **nominal Doppler beam direction axis**. The tests are most conveniently carried out at the same time as the sample volume lateral response tests in 5.5.1.2. During these tests, the co-ordinates of the points at which the Doppler power falls to -6 dB of the maximum are calculated and the co-ordinates of their mid point along the two lateral beam axes are noted. These mid-point co-ordinates are those of the measured beam axis. The difference between these co-ordinates and the co-ordinates of the beam position at that depth (see figure 1). It is clearly necessary for the test rig to be such that the co-ordinates of the string and the nominal Doppler beam direction are known within the co-ordinate system defined by the calibrated probe positioning system (see 5.1.4).





Key

A Probe B String SE Side elevation FE Front elevation Dashed line – nominal Doppler beam axis direction

Dotted line - true Doppler beam axis direction

NOTE This figure shows views along the line perpendicular to the plane containing string and nominal Doppler beam direction axis (and along the direction of the string respectively) for beam deviation measurements (e_2 , e_1) at depth d along second (a) and first (b) lateral beam axes.

5.8 Intrinsic broadening

In this test, the Doppler signal is passed to a spectrum analyser capable of displaying spectra acquired for signals of several seconds duration and allowing measurement of the **6 dB spectral width**. The accuracy of measurement is dependent on the frequency and amplitude resolution of the spectrum analyser, the width of the spectrum being measured and the product of signal duration and frequency resolution. The spectrum analyser, its settings and the signal duration should be such that the **6 dB spectral width** can be measured to an accuracy of 10 %.

The test should be carried out at the working **sample volume** length setting at depths of $0.2kL_{max}$ where k = 1, 2, 3, 4, or as specified by the manufacturer.

This test may be carried out using a string test object as follows. With the Doppler beam direction axis set at the working **Doppler angle** with respect to the string, and the string speed set such that the mean frequency of the Doppler signal spectrum is the Doppler test frequency, the probe is moved axially and laterally while monitoring the Doppler signal power until this is a maximum. The **6 dB spectral width** of the measured Doppler signal spectrum is expressed as a percentage of the test frequency. This figure is reported as the percentage intrinsic spectral broadening.

5.9 Dead zone

Using a string **Doppler test object**, the **target depth** is decreased until the Doppler signal power reduces to -20 dB of the maximum power recorded in 5.4.1. The **target depth** is then reported as the **dead zone** thickness.

5.10 Acoustic working frequency

This should be measured by the zero-crossing frequency or spectrum analysis methods as detailed in 3.4 of IEC 61102.

5.11 Flow direction separation

This should be measured as detailed in 2.6 of IEC 61206.

5.12 Velocity estimation accuracy

These tests are to be carried out on a **system** having a velocity estimation facility.

Using a string **Doppler test object** under the standard working conditions, the velocity estimated by the **Doppler ultrasound system** is compared with the true string speed. The error is reported for string speeds from -2 m/s to +2 m/s in steps of 0,04 m/s, or speeds corresponding to ± 0.8 of the frequency at which **aliasing** occurs, whichever leads to the smallest **Doppler frequency** magnitude.

Using a fixed string speed giving a mean Doppler shift equal to the test frequency at a **Doppler** angle of 60°, the velocity measurement error is measured and reported for **Doppler angles** at extremes for the particular **system** under test and for a number of intermediate angles.

5.13 Volume flow estimation accuracy

This test should be carried out on systems having a flow estimation facility. The test should be carried out as detailed in 2.7 of IEC 61206. A suitable flow test object is described in [10].

5.14 Maximum, mean, mode and median frequency estimation accuracy

The ability of the Doppler system to estimate correctly various spectral parameters, such as maximum Doppler frequency, mean frequency, etc., from the detected ultrasound signal depends on a number of machine-dependent factors, such as the algorithms used for Doppler detection and spectral analysis, and also on the shape and width of the Doppler spectrum. The accuracy of estimation of spectral parameters will be dependent on their rate of change. The tests below aim to assess the estimation accuracy of various spectral parameters, and the dependence of accuracy on the rate of change of these parameters. These tests are only concerned with the effect of machine processing on accuracy, and are not concerned with physical effects, such as geometric spectral broadening.

These tests require the generation of a Doppler signal with a time-varying spectrum of controllable and known frequency content. The only test object which meets these criteria is an electronic one.

Firstly, the test object should be used with a simulated Doppler signal having a constant spectral shape and constant mean frequency. At least three spectra should be used: a narrow symmetrical spectrum having a **20 dB spectral width** in the range 0,1 to 0,2 times the spectral mean frequency, a wide symmetrical spectrum having a **20 dB spectral width** in the range 0,4 to 0,6 times the spectral mean frequency and a wide asymmetrical spectrum having a **20 dB spectral width** in the range 0,4 to 0,6 times the spectral mean frequency and a mode frequency greater than the mean frequency by at least 0,2 times the spectral mean frequency. For each spectrum, tests should be carried out using three positive and three negative mean frequencies. In each case, the three frequencies should be such that

- the minimum frequency in the spectrum is equal to the wall-thump filter 3 dB cut-off frequency;
- the maximum frequency in the spectrum is equal to the upper 3 dB cut-off of the system frequency response (see 5.4.1);
- they are mid-way between these two frequencies.

Secondly, the system should be tested using simulated Doppler signals with sinusoidal meanfrequency variation. The amplitude of the mean-frequency variation should be such that the peak frequency generated is equal to the upper 3 dB cut-off of the system frequency response. For each spectrum shape, the frequency f_{vm} of the sinusoidal mean-frequency variation should be varied from 1,0 Hz to 50 Hz in at least six steps and the error in the calculated maximum, mean, mode and median frequencies noted. Typical spectra and minimum/mean/maximum frequency variations are shown in figures 2 and 3. The error for each f_{vm} /spectrum combination should be quoted as the average bias and standard deviation over a cycle calculated from a sufficiently large number of cycles such that the estimation error for these two quantities is less than 10 %. The above tests should be carried out at signal-to-noise ratios of 0 dB, 10 dB, 20 dB and 40 dB.



- 24 -

Figure 2 – Sinusoidal variation of minimum, mean and maximum simulated Doppler signal frequencies



NOTE This figure shows a narrow positive symmetrical spectrum (a), a wide positive symmetrical spectrum (b), a wide positive asymmetrical spectrum (c) and a wide negative asymmetrical spectrum (d). Frequencies indicated are minimum (1), mean (2), maximum (3) and mode (4). The mode and mean frequencies are equal for the symmetrical spectra.

Figure 3 – Simulated Doppler spectra with a constant mean frequency or at an instant when using a sinusoidal mean-frequency variation

5.15 Velocity waveform indices estimation accuracy

Doppler ultrasound systems often incorporate algorithms to compute indices of velocity waveform shape such as pulsatility index, resistance index and A/B ratio [4]. These may be computed from the maximum frequency envelope of the displayed spectrum automatically or from a trace of the envelope performed by the operator. The results are dependent not only on the algorithms used but also on the effect of the sample volume and spectral analysis method, since these may give rise to maximum frequency waveform shapes which are different from the shape of the velocity waveforms. Ideally, estimation accuracy of the velocity waveform indices should be assessed by the use of a pulsatile flow test object with an accurately known time-varying velocity field and having a negligible distorting effect on the sample volume of the system. In the absence of such a test object, the algorithms used to calculate the indices may be tested, in isolation, by the use of an electronic test object delivering simulated Doppler signals of known time-varying spectra. Since the number and type of indices calculated by Doppler ultrasound systems are very variable and there is no consensus on the required accuracy, it is sufficient that tests are performed to test the estimation accuracy declared by the manufacturer.

When using a pulsatile flow-test object, tests should be carried out using waveforms of known index values covering the declared range for the system under test. The centre of the test object vessel should be at the working depth, and the angle between the vessel axis and the nominal Doppler beam direction axis should be at the working Doppler angle. The error between the measured index and the known index of the pulsatile flow should be declared for each waveform tested. The test should be carried out at signal-to-noise ratios of 0 dB, 10 dB, 20 dB and 40 dB, and the ratio may be adjusted by the use of the system's transmitted output power control. The test should be carried out with both a narrow spectrum achieved by the use of a vessel of diameter greater than five times the greater of the 6 dB width or length of the smaller of the 6 dB width or length of the sample volume.

When using an electronic test object, Doppler signals of known maximum frequency waveforms should be generated and the system should be tested with narrow and wide spectra. The spectra used for the frequency estimation tests in 5.14 would be suitable. Since there is also no consensus on what constitutes "maximum" frequency (a typical Doppler spectrum gradually reduces to zero as a consequence of intrinsic spectral broadening and window broadening, for example, and the true maximum is not used), the definition of the maximum frequency of the simulated Doppler signal spectrum (for example, the frequency below which the spectrum contains 95 % of the total Doppler signal power) should be declared along with the results of these tests. The range of velocity indices and signal-to-noise ratios should be as for the tests using the pulsatile flow test object.

In order to test the response of the system to the varying amplitude velocity waveforms obtained in practice as result of breathing or other involuntary patient movement, the electronic test object should be capable of generating simulated Doppler signals for which the velocity waveform amplitude reduces from its maximum value randomly and smoothly to a minimum value of half the maximum amplitude so that this change from an amplitude maximum to a minimum takes place over a minimum time of two and a maximum time of 10 cardiac cycles. The deviation of each index calculated by the system from the index of the maximum amplitude waveform should be declared.

The above tests should be carried out with cardiac periods of 400 ms, 800 ms and 1 600 ms or at three periods spanning the declared range of the system.

6 Doppler test objects

6.1 Test objects

Test objects suitable for pulsed **Doppler ultrasound systems** are described briefly in Section 3 of IEC 61206 and in [10]. In addition, an electronic **Doppler test object** is described briefly here.

Flow, string, belt and electronic test objects together with their strengths and limitations are described in [2].

It should be noted that the moving belt test object consisting of a distributed target in the form of a motor driven belt of scattering material in a fluid bath [5] is suitable for some of the Doppler frequency response (5.4.1, 5.4.2, 5.4.3.2) tests, the intrinsic broadening test (5.8) and the velocity estimation accuracy test (5.12).

It is important that those characteristics of the test object used which are known or are likely to have an effect on the test results should be reported along with the test results.

6.2 Electronic test object



Figure 4a – Electronic test object



Figure 4b – Alternative coupling arrangement between probe and test object transducers

In an electronic, acoustically coupled test-object, transmitted pulses from the probe of the system under test received by one transducer of the test object are mixed with a simulated Doppler signal to generate Doppler shifted pulses which are retransmitted back to the probe. The system under test thus receives pulses as if from targets generating the Doppler shifts of the simulated Doppler signal. Mixing by simple multiplication leads to a received r.f. signal containing two sidebands and therefore Doppler signals simulating simultaneous forward and receding flow [6]. A single sideband system with the facility allowing the operator to switch to either one or the other sideband has been described by Wallace and Whittingham [7]. The system shown in figure 4a will allow the generation of a Doppler signal without such constraints, that is, with any time varying frequency content within the Nyquist bounds (± 0.5 PRF). The pulse received on transducer 1 is phase-shifted by 90° and the original and phaseshifted signals applied to the real and imaginary connections of one input of a complex multiplier Xc. The other input is supplied by the in-phase and quadrature components of a simulated Doppler signal. The real (or imaginary) part of the complex multiplier provides the Doppler shifted signal to be transmitted (from transducer 2) back to the probe of the system under test. Coupling of the transducers of the test object to the probe can be via a perspex block as shown and as used by the authors above, or possibly by an angled partial reflector as shown in figure 4b. The simulated Doppler signal may vary from a single-frequency sine wave to a simulated Doppler signal having the time-varying spectrum typical of that obtained from arteries. This could be generated directly in analogue form, using a noise source feeding a filter with a time-varying frequency response [8], or in digital form [9] followed by a digital-toanalogue converter. In order to assess the response of the system under test to clutter signals, higher amplitude Doppler signals may be added to those simulating those from blood. Clearly, the Doppler signal simulator should generate spurious signals at a much lower level than those anticipated from the system under test.

- 28 -

Annex A (normative)

Description of pulsed Doppler ultrasound systems

A.1 Single-channel system

A continuous-wave **Doppler ultrasound system** is described in annex A of IEC 61206. The description here concentrates on the differences between that and a pulsed **Doppler ultrasound system**. Reference [4] provides additional details of instrumentation design.

Unlike the continuous-wave system, the pulsed **Doppler ultrasound system** has range resolution. It achieves this by periodically transmitting a pulse consisting of a few cycles of ultrasound, rather than transmitting continuously and sampling the output of the **Doppler demodulator** after a delay following each pulsed transmission. The system is sensitive to flow only within a small **sample volume** (sv) at a distance from the transducer such that the transit time of the ultrasound pulse to the **sample volume** and back is equal to the delay between transmission and receiver sampling.

The block diagram of such a **system** is shown in figure A.1. Since, unlike the continuous wave system, the functions of transmission and reception are separated in time, a single transducer may be used as shown. The transmitter gate is opened periodically at the **pulse repetition** frequency of the system. The duration for which the gate is open together with the filtering action of the transducer determines the transmitted pulse length. After a delay the output of the **Doppler demodulator** is sampled and held until the next sample is taken. In some systems, the output of the **Doppler demodulator** is averaged during the time that the **range gate** (sample) is open. In these, the **sample volume** length is determined primarily by the transmitted pulse length and the **range gate** duration. In others, the **range gate** is open for a time short compared to the time constant of the low-pass filter following the multiplier in the **Doppler demodulator**, and in this case the transmitted pulse length together with this filter time constant determine the **sample volume** length. The **sample volume** width is determined by the width of the ultrasound beam. The **sample volume** depth is varied by altering the delay between transmission and **range-gate**.

The signal from the sample-and-hold circuit is amplified and filtered to give the signal at the **Doppler output connector**. The filters incorporate a wall-thump filter as in continuous wave systems. The low-pass filtration of the signal from the **Doppler demodulator** is determined by the **pulse repetition frequency** of the system. Since the system essentially samples the Doppler signal from a **sample volume**, the spectrum of the signal from the sample-and-hold consists of the required signal spectrum together with copies occurring at multiples of the **pulse repetition frequency**. It is necessary to set the cut-off frequency of the low-pass filter such that frequency components greater than half the **pulse repetition frequency** are strongly attenuated. The maximum **Doppler frequency** is therefore limited to half the **pulse repetition** frequency. As in the continuous wave instrument a flow direction discrimination can be achieved by using an offset frequency reference or by the use of two reference signals with a 90° phase shift. The **system** may incorporate a means of listening to the Doppler signal, such as a loudspeaker or headphones, and a means of frequency analysis such as a spectrum analyser.

A.2 Multi-channel system

The system described is called a single-channel pulsed Doppler system. A multi-channel system, having a number of **sample volumes** operating simultaneously, which can monitor the flow at a number of positions along the beam at the same time, is constructed by having a number of **Doppler demodulator** sampling gates together with their own amplifiers and filters, each gate being opened at a different time following transmission.

A.3 Aliasing

A problem which is peculiar to pulsed Doppler systems is that of **aliasing**. Signals need to be sampled at least twice per cycle of their highest frequency component in order to resolve unambiguously that component. If the sampling frequency is too low, then the frequency of the sampled signal is in error. In a **non-directional system**, the indicated **Doppler frequency** will be the true **Doppler frequency** mirrored in the **Nyquist limit frequency**. In a **directional system**, the indicated frequency of the Doppler signal is the true **Doppler frequency** mirrored in sign. In most systems the Doppler signal may be frequency shifted (a **baseline frequency shift**) before analysis so that zero target velocity corresponds to the **baseline frequency shift** rather than zero frequency. For targets having unequal positive and negative velocity excursions this allows the operator to make maximum use of the available frequency range (for example the range ±0,5 PRF can be shifted to the range -0,2 PRF to 0,8 PRF).

A.4 Duplex and triplex scanners

The combination of a pulsed **Doppler ultrasound system** and a real-time ultrasonic **B-mode** imager is known as a **duplex scanner**. In these scanners the Doppler ultrasound beam and **sample volume** are shown displayed on the **B-mode** image allowing the operator to position the **sample volume** at selected sites within imaged blood vessels. **Duplex scanners** usually have a means of measuring the **Doppler angle** by means of an electronically generated marker line at the position of the **sample volume** which may be rotated by the operator until it is lined up with the axis of the blood vessel. In this case the system can calculate the angle between the beam and vessel axes, use this to convert from measured **Doppler frequency** to blood velocity and display velocity rather than frequency information. In addition, flow may be measured by incorporating a measurement of a blood vessel cross-sectional area using the electronic callipers of the **B-mode** system. Usually, measurements are made by measuring a diameter of the blood vessel and assuming a circular cross-section.

The combination of a **duplex scanner** and a colour flow imager is known as a **triplex scanner**.



- 30 -

Figure A.1 – Single-channel, non-directional pulsed Doppler ultrasound system



- 31 -

Key

- a1 First lateral axis
- a2 Second lateral axis
- a3 Third lateral axis

Figure A.2 – Beam axes for Doppler beam from linear array probe

Bibliography

- [1] AIUM Standards Committee, *Performance criteria and measurements for Doppler ultrasound devices,* American Institute for Ultrasound in Medicine, USA, 1993.
- [2] Hoskins, P. R., Sherriff, S. B. and Evans, J. A., *Testing of Doppler ultrasound equipment*, Institute of Physical Sciences in Medicine, York, UK, 1994.
- [3] IEC 61266:1994, Ultrasonics Hand-held probe Doppler foetal heartbeat detectors Performance requirements and methods of measurement and reporting
- [4] Evans, D. H., McDicken, W. N., Skidmore, R. and Woodcock J. P., *Doppler ultrasound: physics, instrumentation and clinical applications*, Wiley, Chichester, UK, 1989.
- [5] Rickey, D. W., Rankin, R., and Fenster, A. "A velocity evaluation phantom for colour flow and pulsed Doppler instruments", *Ultrasound Med. Biol.*, vol. 18, pp. 479-494, 1992.
- [6] Evans, J. A., Price, R. and Luhana, F. "A novel testing device for Doppler ultrasound equipment", *Phys. Med. Biol.*, vol. 34, pp. 1701-1702, 1989.
- [7] Wallace, J. J. A., Martin, K. and Whittingham, T. A. "An experimental single-sideband acoustical re-injection test method for Doppler systems", *Physiol. Meas.*, vol. 14, pp. 479-484, 1993.
- [8] Bastos, C. A. C. and Fish, P. J. "A Doppler signal simulator", *Clin. Phys. Physiol. Meas.* vol. 12, pp. 177-183, 1991.
- [9] Wang, Y-Y. and Fish, P. J. "Simulating nonstationary in-phase and quadrature Doppler signals using a time-varying impulse response filter", *Ultrasound Med. Biol.*, vol. 22, pp. 529-531, 1996.
- [10] IEC 61685, Ultrasonics Flow measurement systems Flow test object (under consideration).



The IEC would like to offer you the best quality standards possible. To make sure that we continue to meet your needs, your feedback is essential. Would you please take a minute to answer the questions overleaf and fax them to us at +41 22 919 03 00 or mail them to the address below. Thank you!

Customer Service Centre (CSC)

International Electrotechnical Commission 3, rue de Varembé 1211 Genève 20 Switzerland

or

Fax to: IEC/CSC at +41 22 919 03 00

Thank you for your contribution to the standards-making process.



Nicht frankieren Ne pas affranchir



Non affrancare No stamp required

RÉPONSE PAYÉE

SUISSE

Customer Service Centre (CSC) International Electrotechnical Commission 3, rue de Varembé 1211 GENEVA 20 Switzerland

Q1	Please report on ONE STANDARD ar ONE STANDARD ONLY . Enter the expumber of the standard: (e.g. 60601-	nd (act 1-1)	Q6	If you ticked NOT AT ALL in Question the reason is: <i>(tick all that apply)</i>	ז 1 5
		,		standard is out of date	
				standard is incomplete	
				standard is too academic	
Q2	Please tell us in what capacity(ies) yo	u		standard is too superficial	
	bought the standard (tick all that appl	y).		title is misleading	
				I made the wrong choice	
	purchasing agent			other	
	librarian				
	researcher				
	design engineer		07	Please assess the standard in the	
	safety engineer		Q 1	following categories, using	
	testing engineer			the numbers:	
	marketing specialist			(1) unacceptable,	
	other			(2) below average, (3) average	
				(4) above average,	
03	l work for/in/as a:			(5) exceptional,	
40	(tick all that apply)			(6) not applicable	
				timeliness	
	manufacturing			quality of writing	
	consultant			technical contents	
	government			logic of arrangement of contents	
	test/certification facility			tables, charts, graphs, figures	
	public utility			other	
	education				
	military				
	other		Q8	I read/use the: (tick one)	
04	This standard will be used for:			French text only	
47	(tick all that apply)			English text only	
				both English and Erench texts	
	general reference				
	product research				
	product design/development				
	specifications		Q9	Please share any comment on any	
	tenders			aspect of the IEC that you would like	
	quality assessment				
	certification				
	technical documentation				
	thesis				••••
	manufacturing	L			••••
	other				••••
Q5	This standard meets my needs:				••••
	(tick one)				••••
	not at all				
	nearly				
	fairly well				
	exactly	_			

I

I



La CEI ambitionne de vous offrir les meilleures normes possibles. Pour nous assurer que nous continuons à répondre à votre attente, nous avons besoin de quelques renseignements de votre part. Nous vous demandons simplement de consacrer un instant pour répondre au questionnaire ci-après et de nous le retourner par fax au +41 22 919 03 00 ou par courrier à l'adresse ci-dessous. Merci !

Centre du Service Clientèle (CSC)

Commission Electrotechnique Internationale 3, rue de Varembé 1211 Genève 20 Suisse

ou

Télécopie: CEI/CSC +41 22 919 03 00

Nous vous remercions de la contribution que vous voudrez bien apporter ainsi à la Normalisation Internationale.



Veuillez ne mentionner qu' UNE SEUL NORME et indiquer son numéro exac	. E t:	Q5	Cette norm <i>(une seule</i>
			pas du tout à peu près assez bien parfaitemer
En tant qu'acheteur de cette norme, quelle est votre fonction? <i>(cochez tout ce qui convient)</i> Je suis le/un: agent d'un service d'achat		Q6	Si vous ave Q5, c'est pe <i>(cochez tou</i>
bibliothécaire chercheur ingénieur concepteur ingénieur sécurité ingénieur d'essais spécialiste en marketing autre(s)			la norme es la norme es la norme es le titre est e je n'ai pas autre(s)
Je travaille: (cochez tout ce qui convient)		Q7	Veuillez éva dessous en (1) inaccep (2) au-dess
dans l'industrie comme consultant pour un gouvernement			(3) moyen,(4) au-dess(5) exception(6) sans ob
pour un organisme d'essais/ certification dans un service public dans l'enseignement			publication qualité de l contenu teo disposition
comme militaire autre(s)			tableaux, d figures autre(s)
Cette norme sera utilisée pour/commo (cochez tout ce qui convient)	e	Q8	Je lis/utilise uniquemen
ouvrage de référence une recherche de produit une étude/développement de produit			les textes a
des spécifications des soumissions une évaluation de la qualité		Q9	Veuillez no observatior
une certification une documentation technique			
une thèse la fabrication			

autre(s).....

Q1

Q2

Q3

Q4

e répond-elle à vos besoins: réponse)

pas du tout	
à peu près	
assez bien	
parfaitement	

ez répondu PAS DU TOUT à our la/les raison(s) suivantes: ut ce qui convient)

la norme a besoin d'être révisé	e 🛛
la norme est incomplète	
la norme est trop théorique	
la norme est trop superficielle	
le titre est équivoque	
je n'ai pas fait le bon choix	
autre(s)	

Veuillez évaluer chacun des critères dessous en utilisant les chiffres (1) inacceptable, (2) au-dessous de la moyenne, (3) moyen, (4) au-dessus de la moyenne, (5) exceptionnel, (6) sans objet	ci-
publication en temps opportun qualité de la rédaction contenu technique disposition logique du contenu tableaux, diagrammes, graphiques, figures autre(s)	·····

- e: (une seule réponse) it le texte français nt le texte anglais anglais et français
- us faire part de vos ns éventuelles sur la CEI:

LICENSED TO MECON Limited. - RANCHI/BANGALORE FOR INTERNAL USE AT THIS LOCATION ONLY, SUPPLIED BY BOOK SUPPLY BUREAU.



ICS 17.140.50