



INTERNATIONAL STANDARD ISO 2865 : 1973
TECHNICAL CORRIGENDUM 1

Published 1991-07-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION - МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ - ORGANISATION INTERNATIONALE DE NORMALISATION

Aluminium oxide primarily used for the production of aluminium – Determination of boron content – Curcumin spectrophotometric method

TECHNICAL CORRIGENDUM 1

Oxyde d'aluminium principalement utilisé pour la production de l'aluminium – Dosage du bore – Méthode spectrophotométrique à la curcumine

RECTIFICATIF TECHNIQUE 1

Technical corrigendum 1 to International Standard ISO 2865:1973 was prepared by Technical Committee ISO/TC 47, *Chemistry*, Sub-Committee SC 7, *Aluminium oxide, cryolite, aluminium fluoride, sodium fluoride, carbonaceous products for the aluminium industry*.

Page 1

Clause 2

Replace the reference to ISO/R 802 by a reference to ISO 802:1976.

Add the following reference:

"ISO 2927:1973, *Aluminium oxide used primarily for the production of aluminium – Sampling.*"

Page 2

Subclause 6.1

Replace the text of subclause 6.1 by the following text:

"Weigh, to the nearest 0,001 g, 1 g of the laboratory sample, dried at 300 °C, prepared as described in ISO 2927 and ISO 802:1976, subclause 3.3."

UDC 661.862.22 : 546.27 : 543.42

Ref. No. ISO 2865 : 1973/Cor.1 : 1991 (E)

Descriptors : aluminium, aluminium oxide, chemical analysis, determination of content, boron, spectrometric method.

© ISO 1991

Printed in Switzerland

This page intentionally left blank